

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE042619\
 Data File : BE099386.D
 Acq On : 26 Apr 2019 19:55
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 27 06:51:50 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.560	0.490	12.5	86	0.00
3	n-Nitrosodimethylamine	0.301	0.297	1.3	113	-0.01
4 S	2-Fluorophenol	0.977	1.183	-21.1	118	0.01
5 S	Phenol-d6	1.395	1.725	-23.7	121	0.00
6	bis(2-Chloroethyl)ether	1.301	1.279	1.7	92	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.01
8 S	Nitrobenzene-d5	0.311	0.388	-24.8	129	0.00
9	Naphthalene	4.364	4.753	-8.9	106	0.00
10	Hexachlorobutadiene	0.277	0.301	-8.7	105	0.00
11 SURR	2-Methylnaphthalene-d10	0.706	0.718	-1.7	100	0.01
12	2-Methylnaphthalene	0.743	0.806	-8.5	108	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.133	0.233	-75.2#	167#	0.00
15 S	2-Fluorobiphenyl	1.572	1.660	-5.6	98	0.00
16	Acenaphthylene	1.808	2.055	-13.7	112	0.00
17	Acenaphthene	1.114	1.166	-4.7	99	0.01
18	Fluorene	1.601	1.622	-1.3	95	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.01
20	4-Bromophenyl-phenylether	0.222	0.251	-13.1	97	0.00
21	Hexachlorobenzene	0.213	0.242	-13.6	95	0.00
22	Pentachlorophenol	0.057	0.113	-98.2#	206#	0.00
23	Phenanthrene	1.059	1.077	-1.7	87	0.01
24	Anthracene	0.942	1.025	-8.8	94	0.00
25 SURR	Fluoranthene-d10	4.806	5.145	-7.1	93	0.00
26	Fluoranthene	1.418	1.473	-3.9	90	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
28	Pyrene	1.178	1.194	-1.4	92	0.00
29 S	Terphenyl-d14	0.787	0.817	-3.8	95	0.00
30	Benzo(a)anthracene	1.176	1.325	-12.7	110	0.01
31	Chrysene	1.251	1.276	-2.0	92	0.00
32	Bis(2-ethylhexyl)phthalate	1.070	2.342	-118.9#	239#	0.00
33	Indeno(1,2,3-cd)pyrene	1.236	1.258	-1.8	95	0.03
34 I	Perylene-d12	1.000	1.000	0.0	106	0.01
35	Benzo(b)fluoranthene	1.225	1.204	1.7	104	0.01
36	Benzo(k)fluoranthene	1.221	1.131	7.4	96	0.01
37 C	Benzo(a)pyrene	1.097	1.114	-1.5	107	0.01
38	Dibenzo(a,h)anthracene	1.121	1.001	10.7	94	0.02
39	Benzo(g,h,i)perylene	1.137	0.895	21.3	83	0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE042619\
Data File : BE099386.D
Acq On : 26 Apr 2019 19:55
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 27 06:51:50 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		