

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE042619\
 Data File : BE099388.D
 Acq On : 26 Apr 2019 21:43
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 27 06:54:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.560	0.448	20.0	80	0.00
3	n-Nitrosodimethylamine	0.301	0.295	2.0	115	-0.01
4 S	2-Fluorophenol	0.977	1.193	-22.1	121	0.01
5 S	Phenol-d6	1.395	1.728	-23.9	124	0.00
6	bis(2-Chloroethyl)ether	1.301	1.298	0.2	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.01
8 S	Nitrobenzene-d5	0.311	0.382	-22.8	134	0.00
9	Naphthalene	4.364	4.699	-7.7	110	0.00
10	Hexachlorobutadiene	0.277	0.306	-10.5	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.706	0.715	-1.3	105	0.01
12	2-Methylnaphthalene	0.743	0.800	-7.7	112	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.133	0.217	-63.2#	167#	0.00
15 S	2-Fluorobiphenyl	1.572	1.587	-1.0	101	0.00
16	Acenaphthylene	1.808	1.958	-8.3	115	0.00
17	Acenaphthene	1.114	1.111	0.3	102	0.01
18	Fluorene	1.601	1.547	3.4	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
20	4-Bromophenyl-phenylether	0.222	0.247	-11.3	100	0.00
21	Hexachlorobenzene	0.213	0.231	-8.5	95	0.00
22	Pentachlorophenol	0.057	0.102	-78.9#	193#	0.00
23	Phenanthrene	1.059	1.059	0.0	89	0.00
24	Anthracene	0.942	1.020	-8.3	98	0.00
25 SURR	Fluoranthene-d10	4.806	4.994	-3.9	94	0.00
26	Fluoranthene	1.418	1.438	-1.4	92	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
28	Pyrene	1.178	1.196	-1.5	94	0.00
29 S	Terphenyl-d14	0.787	0.810	-2.9	96	0.00
30	Benzo(a)anthracene	1.176	1.316	-11.9	110	0.01
31	Chrysene	1.251	1.252	-0.1	91	0.00
32	Bis(2-ethylhexyl)phthalate	1.070	2.304	-115.3#	238#	0.00
33	Indeno(1,2,3-cd)pyrene	1.236	1.284	-3.9	98	0.03
34 I	Perylene-d12	1.000	1.000	0.0	106	0.01
35	Benzo(b)fluoranthene	1.225	1.196	2.4	104	0.01
36	Benzo(k)fluoranthene	1.221	1.179	3.4	100	0.01
37 C	Benzo(a)pyrene	1.097	1.126	-2.6	109	0.01
38	Dibenzo(a,h)anthracene	1.121	1.032	7.9	98	0.02
39	Benzo(g,h,i)perylene	1.137	0.883	22.3	82	0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE042619\
Data File : BE099388.D
Acq On : 26 Apr 2019 21:43
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 27 06:54:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		