

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE042619\
 Data File : BE099388.D
 Acq On : 26 Apr 2019 21:43
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 27 06:54:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2	1,4-Dioxane	0.400	0.320	20.0	80	0.00
3	n-Nitrosodimethylamine	0.400	0.391	2.3	115	-0.01
4 S	2-Fluorophenol	0.400	0.488	-22.0	121	0.01
5 S	Phenol-d6	0.400	0.496	-24.0	124	0.00
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	95	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	103	0.01
8 S	Nitrobenzene-d5	0.400	0.492	-23.0	134	0.00
9	Naphthalene	0.400	0.431	-7.7	110	0.00
10	Hexachlorobutadiene	0.400	0.442	-10.5	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.405	-1.3	105	0.01
12	2-Methylnaphthalene	0.400	0.431	-7.7	112	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.400	0.653	-63.2#	167	0.00
15 S	2-Fluorobiphenyl	0.400	0.404	-1.0	101	0.00
16	Acenaphthylene	0.400	0.433	-8.2	115	0.00
17	Acenaphthene	0.400	0.399	0.3	102	0.01
18	Fluorene	0.400	0.386	3.5	98	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	90	0.00
20	4-Bromophenyl-phenylether	0.400	0.446	-11.5	100	0.00
21	Hexachlorobenzene	0.400	0.434	-8.5	95	0.00
22	Pentachlorophenol	0.400	0.693	-73.2#	193	0.00
23	Phenanthrene	0.400	0.400	0.0	89	0.00
24	Anthracene	0.400	0.433	-8.2	98	0.00
25 SURR	Fluoranthene-d10	0.400	0.416	-4.0	94	0.00
26	Fluoranthene	0.400	0.406	-1.5	92	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	94	0.00
28	Pyrene	0.400	0.406	-1.5	94	0.00
29 S	Terphenyl-d14	0.400	0.411	-2.7	96	0.00
30	Benzo(a)anthracene	0.400	0.448	-12.0	110	0.01
31	Chrysene	0.400	0.400	0.0	91	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.861	-115.2#	238	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.416	-4.0	98	0.03
34 I	Perylene-d12	0.400	0.400	0.0	106	0.01
35	Benzo(b)fluoranthene	0.400	0.391	2.3	104	0.01
36	Benzo(k)fluoranthene	0.400	0.386	3.5	100	0.01
37 C	Benzo(a)pyrene	0.400	0.410	-2.5	109	0.01
38	Dibenzo(a,h)anthracene	0.400	0.368	8.0	98	0.02
39	Benzo(g,h,i)perylene	0.400	0.310	22.5	82	0.03

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE042619\
Data File : BE099388.D
Acq On : 26 Apr 2019 21:43
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 27 06:54:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				