

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE042619\
 Data File : BE099405.D
 Acq On : 27 Apr 2019 7:55
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 29 01:56:21 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2	1,4-Dioxane	0.560	0.443	20.9	71	0.00
3	n-Nitrosodimethylamine	0.301	0.277	8.0	97	-0.01
4 S	2-Fluorophenol	0.977	1.137	-16.4	103	0.01
5 S	Phenol-d6	1.395	1.649	-18.2	106	0.00
6	bis(2-Chloroethyl)ether	1.301	1.208	7.1	79	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	87	0.01
8 S	Nitrobenzene-d5	0.311	0.384	-23.5	114	0.00
9	Naphthalene	4.364	4.715	-8.0	94	0.00
10	Hexachlorobutadiene	0.277	0.308	-11.2	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.706	0.698	1.1	87	0.01
12	2-Methylnaphthalene	0.743	0.806	-8.5	96	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
14 S	2,4,6-Tribromophenol	0.133	0.236	-77.4#	155#	0.00
15 S	2-Fluorobiphenyl	1.572	1.574	-0.1	86	0.00
16	Acenaphthylene	1.808	1.948	-7.7	97	0.00
17	Acenaphthene	1.114	1.095	1.7	86	0.01
18	Fluorene	1.601	1.569	2.0	85	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	79	0.01
20	4-Bromophenyl-phenylether	0.222	0.234	-5.4	82	0.00
21	Hexachlorobenzene	0.213	0.233	-9.4	83	0.01
22	Pentachlorophenol	0.057	0.104	-82.5#	173#	0.00
23	Phenanthrene	1.059	1.043	1.5	76	0.01
24	Anthracene	0.942	0.978	-3.8	82	0.00
25 SURR	Fluoranthene-d10	4.806	4.935	-2.7	81	0.00
26	Fluoranthene	1.418	1.413	0.4	79	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
28	Pyrene	1.178	1.185	-0.6	82	0.00
29 S	Terphenyl-d14	0.787	0.807	-2.5	84	0.00
30	Benzo(a)anthracene	1.176	1.266	-7.7	93	0.01
31	Chrysene	1.251	1.222	2.3	78	0.00
32	Bis(2-ethylhexyl)phthalate	1.070	2.316	-116.4#	210#	0.00
33	Indeno(1,2,3-cd)pyrene	1.236	1.124	9.1	75	0.03
34 I	Perylene-d12	1.000	1.000	0.0	91	0.02
35	Benzo(b)fluoranthene	1.225	1.200	2.0	90	0.02
36	Benzo(k)fluoranthene	1.221	1.151	5.7	85	0.02
37 C	Benzo(a)pyrene	1.097	1.119	-2.0	93	0.01
38	Dibenzo(a,h)anthracene	1.121	0.942	16.0	77	0.03
39	Benzo(g,h,i)perylene	1.137	0.788	30.7#	63	0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE042619\
Data File : BE099405.D
Acq On : 27 Apr 2019 7:55
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 29 01:56:21 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		