

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE042619\
 Data File : BE099405.D
 Acq On : 27 Apr 2019 7:55
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 29 01:56:21 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	89	0.00
2	1,4-Dioxane	0.400	0.316	21.0	71	0.00
3	n-Nitrosodimethylamine	0.400	0.367	8.3	97	-0.01
4 S	2-Fluorophenol	0.400	0.465	-16.3	103	0.01
5 S	Phenol-d6	0.400	0.473	-18.2	106	0.00
6	bis(2-Chloroethyl)ether	0.400	0.371	7.3	79	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	87	0.01
8 S	Nitrobenzene-d5	0.400	0.494	-23.5	114	0.00
9	Naphthalene	0.400	0.432	-8.0	94	0.00
10	Hexachlorobutadiene	0.400	0.445	-11.2	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.395	1.3	87	0.01
12	2-Methylnaphthalene	0.400	0.434	-8.5	96	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	88	0.00
14 S	2,4,6-Tribromophenol	0.400	0.710	-77.5#	155	0.00
15 S	2-Fluorobiphenyl	0.400	0.400	0.0	86	0.00
16	Acenaphthylene	0.400	0.431	-7.7	97	0.00
17	Acenaphthene	0.400	0.393	1.8	86	0.01
18	Fluorene	0.400	0.392	2.0	85	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	79	0.01
20	4-Bromophenyl-phenylether	0.400	0.423	-5.7	82	0.00
21	Hexachlorobenzene	0.400	0.437	-9.2	83	0.01
22	Pentachlorophenol	0.400	0.705	-76.2#	173	0.00
23	Phenanthrene	0.400	0.394	1.5	76	0.01
24	Anthracene	0.400	0.415	-3.7	82	0.00
25 SURR	Fluoranthene-d10	0.400	0.411	-2.7	81	0.00
26	Fluoranthene	0.400	0.399	0.3	79	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	83	0.00
28	Pyrene	0.400	0.402	-0.5	82	0.00
29 S	Terphenyl-d14	0.400	0.410	-2.5	84	0.00
30	Benzo(a)anthracene	0.400	0.430	-7.5	93	0.01
31	Chrysene	0.400	0.391	2.3	78	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.866	-116.5#	210	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.364	9.0	75	0.03
34 I	Perylene-d12	0.400	0.400	0.0	91	0.02
35	Benzo(b)fluoranthene	0.400	0.392	2.0	90	0.02
36	Benzo(k)fluoranthene	0.400	0.377	5.8	85	0.02
37 C	Benzo(a)pyrene	0.400	0.408	-2.0	93	0.01
38	Dibenzo(a,h)anthracene	0.400	0.336	16.0	77	0.03
39	Benzo(g,h,i)perylene	0.400	0.277	30.8#	63	0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE042619\
Data File : BE099405.D
Acq On : 27 Apr 2019 7:55
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 29 01:56:21 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				