

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050119\
 Data File : BE099454.D
 Acq On : 1 May 2019 9:03
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 02 07:32:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 29 18:36:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.550	0.491	10.7	111	0.00
3	n-Nitrosodimethylamine	0.332	0.303	8.7	110	0.00
4 S	2-Fluorophenol	1.106	1.178	-6.5	119	0.00
5 S	Phenol-d6	1.471	1.830	-24.4	136	0.00
6	bis(2-Chloroethyl)ether	1.269	1.252	1.3	108	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	130	0.00
8 S	Nitrobenzene-d5	0.375	0.368	1.9	127	0.00
9	Naphthalene	4.359	4.496	-3.1	132	0.00
10	Hexachlorobutadiene	0.303	0.300	1.0	126	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.769	-6.7	140	0.00
12	2-Methylnaphthalene	0.758	0.820	-8.2	141	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	140	0.00
14 S	2,4,6-Tribromophenol	0.252	0.258	-2.4	148	0.00
15 S	2-Fluorobiphenyl	1.526	1.515	0.7	136	-0.02
16	Acenaphthylene	1.957	1.986	-1.5	142	0.00
17	Acenaphthene	1.118	1.142	-2.1	144	0.00
18	Fluorene	1.649	1.648	0.1	139	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	133	0.00
20	4-Bromophenyl-phenylether	0.233	0.234	-0.4	132	0.00
21	Hexachlorobenzene	0.222	0.234	-5.4	139	0.00
22	Pentachlorophenol	0.100	0.100	0.0	157#	0.00
23	Phenanthrene	1.032	1.019	1.3	129	0.00
24	Anthracene	0.984	0.964	2.0	131	0.00
25 SURR	Fluoranthene-d10	5.391	4.873	9.6	120	0.00
26	Fluoranthene	1.546	1.419	8.2	121	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
28	Pyrene	1.043	1.302	-24.8	118	0.00
29 S	Terphenyl-d14	0.733	0.887	-21.0	115	0.00
30	Benzo(a)anthracene	1.274	1.467	-15.1	109	-0.01
31	Chrysene	1.247	1.404	-12.6	106	0.00
32	Bis(2-ethylhexyl)phthalate	2.216	2.226	-0.5	99	-0.01
33	Indeno(1,2,3-cd)pyrene	1.474	1.656	-12.3	105	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	104	0.00
35	Benzo(b)fluoranthene	1.153	1.222	-6.0	109	0.00
36	Benzo(k)fluoranthene	1.112	1.144	-2.9	102	0.00
37 C	Benzo(a)pyrene	1.090	1.142	-4.8	106	0.00
38	Dibenzo(a,h)anthracene	1.115	1.147	-2.9	104	0.00
39	Benzo(g,h,i)perylene	1.118	1.165	-4.2	105	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050119\
Data File : BE099454.D
Acq On : 1 May 2019 9:03
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 02 07:32:19 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 29 18:36:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			