

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050119\
 Data File : BE099490.D
 Acq On : 2 May 2019 7:35
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 02 08:35:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 29 18:36:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	-0.01
2	1,4-Dioxane	0.550	0.455	17.3	106	0.00
3	n-Nitrosodimethylamine	0.332	0.280	15.7	105	-0.01
4 S	2-Fluorophenol	1.106	1.162	-5.1	121	0.00
5 S	Phenol-d6	1.471	1.772	-20.5	136	0.00
6	bis(2-Chloroethyl)ether	1.269	1.244	2.0	110	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	126	0.00
8 S	Nitrobenzene-d5	0.375	0.393	-4.8	131	0.00
9	Naphthalene	4.359	4.567	-4.8	129	0.00
10	Hexachlorobutadiene	0.303	0.307	-1.3	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.754	-4.6	133	0.00
12	2-Methylnaphthalene	0.758	0.818	-7.9	136	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	137	0.00
14 S	2,4,6-Tribromophenol	0.252	0.275	-9.1	155#	0.00
15 S	2-Fluorobiphenyl	1.526	1.500	1.7	132	-0.02
16	Acenaphthylene	1.957	1.972	-0.8	138	0.00
17	Acenaphthene	1.118	1.153	-3.1	142	0.00
18	Fluorene	1.649	1.650	-0.1	136	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	129	0.00
20	4-Bromophenyl-phenylether	0.233	0.244	-4.7	134	0.00
21	Hexachlorobenzene	0.222	0.239	-7.7	139	0.00
22	Pentachlorophenol	0.100	0.114	-14.0	175#	0.00
23	Phenanthrene	1.032	1.049	-1.6	130	0.00
24	Anthracene	0.984	1.025	-4.2	135	0.00
25 SURR	Fluoranthene-d10	5.391	5.159	4.3	124	0.00
26	Fluoranthene	1.546	1.481	4.2	123	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
28	Pyrene	1.043	1.333	-27.8#	121	0.00
29 S	Terphenyl-d14	0.733	0.911	-24.3	119	0.00
30	Benzo(a)anthracene	1.274	1.486	-16.6	111	-0.01
31	Chrysene	1.247	1.426	-14.4	108	0.00
32	Bis(2-ethylhexyl)phthalate	2.216	2.386	-7.7	107	0.00
33	Indeno(1,2,3-cd)pyrene	1.474	1.657	-12.4	106	0.00
34 I	Perylene-d12	1.000	1.000	0.0	105	0.00
35	Benzo(b)fluoranthene	1.153	1.206	-4.6	108	0.00
36	Benzo(k)fluoranthene	1.112	1.158	-4.1	104	0.00
37 C	Benzo(a)pyrene	1.090	1.142	-4.8	107	0.00
38	Dibenzo(a,h)anthracene	1.115	1.158	-3.9	106	0.00
39	Benzo(g,h,i)perylene	1.118	1.172	-4.8	106	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050119\
Data File : BE099490.D
Acq On : 2 May 2019 7:35
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 02 08:35:05 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 29 18:36:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		