

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050319\
 Data File : BE099530.D
 Acq On : 3 May 2019 15:10
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: May 04 01:27:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 29 18:36:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	-0.01
2	1,4-Dioxane	0.550	0.600	-9.1	114	-0.01
3	n-Nitrosodimethylamine	0.332	0.360	-8.4	110	-0.01
4 S	2-Fluorophenol	1.106	1.243	-12.4	106	0.00
5 S	Phenol-d6	1.471	1.718	-16.8	107	-0.01
6	bis(2-Chloroethyl)ether	1.269	1.288	-1.5	93	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.375	0.406	-8.3	102	0.00
9	Naphthalene	4.359	4.640	-6.4	99	0.00
10	Hexachlorobutadiene	0.303	0.320	-5.6	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.720	0.1	95	0.00
12	2-Methylnaphthalene	0.758	0.789	-4.1	98	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.252	0.310	-23.0	129	0.00
15 S	2-Fluorobiphenyl	1.526	1.534	-0.5	100	0.00
16	Acenaphthylene	1.957	1.955	0.1	101	0.00
17	Acenaphthene	1.118	1.114	0.4	102	0.00
18	Fluorene	1.649	1.696	-2.9	103	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4-Bromophenyl-phenylether	0.233	0.229	1.7	101	0.00
21	Hexachlorobenzene	0.222	0.231	-4.1	107	0.00
22	Pentachlorophenol	0.100	0.105	-5.0	129	0.01
23	Phenanthrene	1.032	1.051	-1.8	105	0.00
24	Anthracene	0.984	0.984	0.0	105	0.00
25 SURR	Fluoranthene-d10	5.391	5.330	1.1	103	0.00
26	Fluoranthene	1.546	1.575	-1.9	105	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
28	Pyrene	1.043	1.265	-21.3	105	0.00
29 S	Terphenyl-d14	0.733	0.883	-20.5	105	0.00
30	Benzo(a)anthracene	1.274	1.450	-13.8	99	0.00
31	Chrysene	1.247	1.429	-14.6	99	0.00
32	Bis(2-ethylhexyl)phthalate	2.216	2.231	-0.7	91	0.00
33	Indeno(1,2,3-cd)pyrene	1.474	1.693	-14.9	99	0.00
34 I	Perylene-d12	1.000	1.000	0.0	96	0.00
35	Benzo(b)fluoranthene	1.153	1.228	-6.5	100	0.00
36	Benzo(k)fluoranthene	1.112	1.186	-6.7	97	0.00
37 C	Benzo(a)pyrene	1.090	1.169	-7.2	100	0.00
38	Dibenzo(a,h)anthracene	1.115	1.179	-5.7	98	0.01
39	Benzo(g,h,i)perylene	1.118	1.189	-6.4	99	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050319\
Data File : BE099530.D
Acq On : 3 May 2019 15:10
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 04 01:27:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 29 18:36:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			