

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050319\
 Data File : BE099545.D
 Acq On : 4 May 2019 00:20
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 04 01:05:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 29 18:36:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.550	0.477	13.3	101	0.00
3	n-Nitrosodimethylamine	0.332	0.363	-9.3	125	-0.01
4 S	2-Fluorophenol	1.106	1.218	-10.1	116	0.00
5 S	Phenol-d6	1.471	1.793	-21.9	125	0.00
6	bis(2-Chloroethyl)ether	1.269	1.250	1.5	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
8 S	Nitrobenzene-d5	0.375	0.428	-14.1	123	0.00
9	Naphthalene	4.359	4.610	-5.8	113	0.00
10	Hexachlorobutadiene	0.303	0.323	-6.6	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.721	0.747	-3.6	113	0.01
12	2-Methylnaphthalene	0.758	0.796	-5.0	114	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.252	0.298	-18.3	140	0.00
15 S	2-Fluorobiphenyl	1.526	1.533	-0.5	113	0.00
16	Acenaphthylene	1.957	1.986	-1.5	116	0.00
17	Acenaphthene	1.118	1.130	-1.1	116	0.01
18	Fluorene	1.649	1.643	0.4	113	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.01
20	4-Bromophenyl-phenylether	0.233	0.247	-6.0	109	0.00
21	Hexachlorobenzene	0.222	0.241	-8.6	112	0.01
22	Pentachlorophenol	0.100	0.118	-18.0	146	0.01
23	Phenanthrene	1.032	1.037	-0.5	103	0.01
24	Anthracene	0.984	1.048	-6.5	111	0.01
25 SURR	Fluoranthene-d10	5.391	5.050	6.3	97	0.01
26	Fluoranthene	1.546	1.485	3.9	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	85	0.01
28	Pyrene	1.043	1.354	-29.8#	98	0.00
29 S	Terphenyl-d14	0.733	0.956	-30.4#	99	0.00
30	Benzo(a)anthracene	1.274	1.509	-18.4	90	0.00
31	Chrysene	1.247	1.434	-15.0	86	0.01
32	Bis(2-ethylhexyl)phthalate	2.216	2.480	-11.9	88	0.00
33	Indeno(1,2,3-cd)pyrene	1.474	1.477	-0.2	75	0.03
34 I	Perylene-d12	1.000	1.000	0.0	86	0.02
35	Benzo(b)fluoranthene	1.153	1.174	-1.8	86	0.01
36	Benzo(k)fluoranthene	1.112	1.168	-5.0	86	0.00
37 C	Benzo(a)pyrene	1.090	1.133	-3.9	87	0.01
38	Dibenzo(a,h)anthracene	1.115	1.026	8.0	77	0.04
39	Benzo(g,h,i)perylene	1.118	0.848	24.2	63	0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050319\
Data File : BE099545.D
Acq On : 4 May 2019 00:20
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 04 01:05:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 29 18:36:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			