

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050319\
 Data File : BE099545.D
 Acq On : 4 May 2019 00:20
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 04 01:05:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 29 18:36:43 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	101	0.00
2	1,4-Dioxane	0.400	0.347	13.3	101	0.00
3	n-Nitrosodimethylamine	0.400	0.438	-9.5	125	-0.01
4 S	2-Fluorophenol	0.400	0.441	-10.2	116	0.00
5 S	Phenol-d6	0.400	0.487	-21.7	125	0.00
6	bis(2-Chloroethyl)ether	0.400	0.394	1.5	101	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	108	0.00
8 S	Nitrobenzene-d5	0.400	0.456	-14.0	123	0.00
9	Naphthalene	0.400	0.423	-5.7	113	0.00
10	Hexachlorobutadiene	0.400	0.426	-6.5	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.415	-3.7	113	0.01
12	2-Methylnaphthalene	0.400	0.420	-5.0	114	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.400	0.473	-18.2	140	0.00
15 S	2-Fluorobiphenyl	0.400	0.402	-0.5	113	0.00
16	Acenaphthylene	0.400	0.406	-1.5	116	0.00
17	Acenaphthene	0.400	0.404	-1.0	116	0.01
18	Fluorene	0.400	0.399	0.3	113	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	0.01
20	4-Bromophenyl-phenylether	0.400	0.425	-6.2	109	0.00
21	Hexachlorobenzene	0.400	0.433	-8.2	112	0.01
22	Pentachlorophenol	0.400	0.570	-42.5#	146	0.01
23	Phenanthrene	0.400	0.402	-0.5	103	0.01
24	Anthracene	0.400	0.426	-6.5	111	0.01
25 SURR	Fluoranthene-d10	0.400	0.375	6.3	97	0.01
26	Fluoranthene	0.400	0.384	4.0	99	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	85	0.01
28	Pyrene	0.400	0.520	-30.0#	98	0.00
29 S	Terphenyl-d14	0.400	0.522	-30.5#	99	0.00
30	Benzo(a)anthracene	0.400	0.474	-18.5	90	0.00
31	Chrysene	0.400	0.460	-15.0	86	0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.480	-20.0	88	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.401	-0.3	75	0.03
34 I	Perylene-d12	0.400	0.400	0.0	86	0.02
35	Benzo(b)fluoranthene	0.400	0.407	-1.7	86	0.01
36	Benzo(k)fluoranthene	0.400	0.420	-5.0	86	0.00
37 C	Benzo(a)pyrene	0.400	0.416	-4.0	87	0.01
38	Dibenzo(a,h)anthracene	0.400	0.368	8.0	77	0.04
39	Benzo(g,h,i)perylene	0.400	0.304	24.0	63	0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050319\
Data File : BE099545.D
Acq On : 4 May 2019 00:20
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 04 01:05:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 29 18:36:43 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				