


Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE050319\
Data File : BE099529.D
Acq On : 3 May 2019 14:33
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE042919.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Apr 29 18:36:43 2019



AutoFind: Scans 2559, 2560, 2561; Background Corrected with Scan 2552

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	20.7	9116	PASS
68	69	0.00	2	1.1	146	PASS
69	198	0.00	100	31.3	13811	PASS
70	69	0.00	2	0.7	92	PASS
127	198	10	80	33.3	14704	PASS
197	198	0.00	2	0.2	68	PASS
198	198	100	100	100.0	44098	PASS
199	198	5	9	7.0	3092	PASS
275	198	10	60	31.3	13821	PASS
365	198	1	100	4.4	1947	PASS
441	198	0.01	100	16.6	7308	PASS
442	198	50	100	97.9	43172	PASS
443	442	15	24	19.3	8337	PASS