

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE050719\
 Data File : BE099571.D
 Acq On : 7 May 2019 19:14
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 08 02:00:01 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 07 14:10:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2	1,4-Dioxane	0.565	0.551	2.5	143	0.00
3	n-Nitrosodimethylamine	0.374	0.341	8.8	122	0.00
4 S	2-Fluorophenol	1.109	1.174	-5.9	143	0.00
5 S	Phenol-d6	1.459	1.553	-6.4	141	0.00
6	bis(2-Chloroethyl)ether	1.216	1.262	-3.8	141	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
8 S	Nitrobenzene-d5	0.385	0.400	-3.9	140	0.00
9	Naphthalene	4.237	20.852	-392.1#	658#	0.00
10	Hexachlorobutadiene	0.308	0.310	-0.6	135	0.00
11 SURR	2-Methylnaphthalene-d10	0.697	0.687	1.4	133	0.00
12	2-Methylnaphthalene	0.718	3.169	-341.4#	584#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
14 S	2,4,6-Tribromophenol	0.266	0.260	2.3	129	0.00
15 S	2-Fluorobiphenyl	1.516	1.577	-4.0	136	0.00
16	Acenaphthylene	1.930	3.243	-68.0#	216#	0.00
17	Acenaphthene	1.078	1.247	-15.7	147	0.00
18	Fluorene	1.540	2.716	-76.4#	223#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
20	4-Bromophenyl-phenylether	0.227	0.228	-0.4	131	0.00
21	Hexachlorobenzene	0.226	0.230	-1.8	129	0.00
22	Pentachlorophenol	0.093	0.085	8.6	131	0.00
23	Phenanthrene	0.986	2.991	-203.3#	388#	0.00
24	Anthracene	0.948	1.336	-40.9#	185#	0.00
25 SURR	Fluoranthene-d10	5.386	5.436	-0.9	131	0.00
26	Fluoranthene	1.518	2.499	-64.6#	211#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	116	0.00
28	Pvrene	1.072	2.232	-108.2#	244#	0.00
29 S	Terphenyl-d14	0.741	0.860	-16.1	136	0.00
30	Benzo(a)anthracene	1.239	1.810	-46.1#	172#	-0.01
31	Chrysene	1.223	1.692	-38.3#	160#	0.00
32	Bis(2-ethylhexyl)phthalate	1.964	2.323	-18.3	140	-0.01
33	Indeno(1,2,3-cd)pyrene	1.452	1.689	-16.3	136	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.103	1.394	-26.4#	155#	0.00
36	Benzo(k)fluoranthene	1.086	1.300	-19.7	148	0.00
37 C	Benzo(a)pyrene	1.064	1.361	-27.9#	157#	-0.01
38	Dibenzo(a,h)anthracene	1.100	1.185	-7.7	133	0.00
39	Benzo(a,h,i)perylene	1.102	1.244	-12.9	137	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE050719\
Data File : BE099571.D
Acq On : 7 May 2019 19:14
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 08 02:00:01 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 07 14:10:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 1			