

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE050819\
 Data File : BE099573.D
 Acq On : 8 May 2019 8:46
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 08 12:25:00 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 07 14:10:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.565	0.587	-3.9	120	0.00
3	n-Nitrosodimethylamine	0.374	0.374	0.0	105	0.00
4 S	2-Fluorophenol	1.109	1.192	-7.5	114	0.00
5 S	Phenol-d6	1.459	1.549	-6.2	110	0.00
6	bis(2-Chloroethyl)ether	1.216	1.193	1.9	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
8 S	Nitrobenzene-d5	0.385	0.375	2.6	99	0.00
9	Naphthalene	4.237	4.506	-6.3	108	0.00
10	Hexachlorobutadiene	0.308	0.325	-5.5	107	0.00
11 SURR	2-Methylnaphthalene-d10	0.697	0.685	1.7	101	0.00
12	2-Methylnaphthalene	0.718	0.726	-1.1	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	83	0.00
14 S	2,4,6-Tribromophenol	0.266	0.245	7.9	89	0.00
15 S	2-Fluorobiphenyl	1.516	1.611	-6.3	101	0.00
16	Acenaphthylene	1.930	1.975	-2.3	96	0.00
17	Acenaphthene	1.078	1.113	-3.2	96	0.00
18	Fluorene	1.540	1.642	-6.6	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
20	4-Bromophenyl-phenylether	0.227	0.231	-1.8	99	0.00
21	Hexachlorobenzene	0.226	0.229	-1.3	96	0.00
22	Pentachlorophenol	0.093	0.076	18.3	87	0.01
23	Phenanthrene	0.986	1.028	-4.3	99	0.00
24	Anthracene	0.948	0.931	1.8	96	0.00
25 SURR	Fluoranthene-d10	5.386	5.462	-1.4	98	0.00
26	Fluoranthene	1.518	1.618	-6.6	102	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
28	Pvrene	1.072	1.160	-8.2	104	0.00
29 S	Terphenyl-d14	0.741	0.790	-6.6	102	0.00
30	Benzo(a)anthracene	1.239	1.383	-11.6	107	-0.01
31	Chrysene	1.223	1.365	-11.6	106	0.00
32	Bis(2-ethylhexyl)phthalate	1.964	2.063	-5.0	102	-0.01
33	Indeno(1,2,3-cd)pyrene	1.452	1.706	-17.5	113	0.00
34 I	Perylene-d12	1.000	1.000	0.0	97	0.00
35	Benzo(b)fluoranthene	1.103	1.182	-7.2	113	0.00
36	Benzo(k)fluoranthene	1.086	1.186	-9.2	116	0.00
37 C	Benzo(a)pyrene	1.064	1.133	-6.5	113	0.00
38	Dibenzo(a,h)anthracene	1.100	1.167	-6.1	113	0.00
39	Benzo(a,h,i)perylene	1.102	1.181	-7.2	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE050819\
Data File : BE099573.D
Acq On : 8 May 2019 8:46
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 08 12:25:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 07 14:10:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			