

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE050819\
 Data File : BE099579.D
 Acq On : 8 May 2019 17:48
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 08 18:24:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 07 14:10:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.565	0.554	1.9	114	0.00
3	n-Nitrosodimethylamine	0.374	0.352	5.9	99	0.00
4 S	2-Fluorophenol	1.109	1.150	-3.7	110	0.00
5 S	Phenol-d6	1.459	1.574	-7.9	112	0.00
6	bis(2-Chloroethyl)ether	1.216	1.196	1.6	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	90	0.01
8 S	Nitrobenzene-d5	0.385	0.411	-6.8	107	0.01
9	Naphthalene	4.237	4.514	-6.5	106	0.01
10	Hexachlorobutadiene	0.308	0.312	-1.3	101	0.01
11 SURR	2-Methylnaphthalene-d10	0.697	0.699	-0.3	101	0.01
12	2-Methylnaphthalene	0.718	0.756	-5.3	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.266	0.280	-5.3	103	0.00
15 S	2-Fluorobiphenyl	1.516	1.596	-5.3	102	0.00
16	Acenaphthylene	1.930	1.943	-0.7	96	0.01
17	Acenaphthene	1.078	1.106	-2.6	97	0.00
18	Fluorene	1.540	1.594	-3.5	97	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
20	4-Bromophenyl-phenylether	0.227	0.226	0.4	96	0.01
21	Hexachlorobenzene	0.226	0.232	-2.7	96	0.00
22	Pentachlorophenol	0.093	0.117	-25.8#	133	0.00
23	Phenanthrene	0.986	1.036	-5.1	99	0.00
24	Anthracene	0.948	0.978	-3.2	99	0.00
25 SURR	Fluoranthene-d10	5.386	5.430	-0.8	96	0.00
26	Fluoranthene	1.518	1.581	-4.2	98	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
28	Pvrene	1.072	1.205	-12.4	100	0.00
29 S	Terphenyl-d14	0.741	0.836	-12.8	100	0.00
30	Benzo(a)anthracene	1.239	1.418	-14.4	102	0.00
31	Chrysene	1.223	1.385	-13.2	99	0.00
32	Bis(2-ethylhexyl)phthalate	1.964	2.233	-13.7	102	0.00
33	Indeno(1,2,3-cd)pyrene	1.452	1.819	-25.3#	111	0.00
34 I	Perylene-d12	1.000	1.000	0.0	95	0.00
35	Benzo(b)fluoranthene	1.103	1.192	-8.1	112	0.00
36	Benzo(k)fluoranthene	1.086	1.136	-4.6	109	0.00
37 C	Benzo(a)pyrene	1.064	1.140	-7.1	112	0.00
38	Dibenzo(a,h)anthracene	1.100	1.196	-8.7	114	0.00
39	Benzo(a,h,i)perylene	1.102	1.182	-7.3	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE050819\
Data File : BE099579.D
Acq On : 8 May 2019 17:48
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 08 18:24:19 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 07 14:10:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			