

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE050819\
 Data File : BE099579.D
 Acq On : 8 May 2019 17:48
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 08 18:24:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 07 14:10:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	0.00
2	1,4-Dioxane	0.400	0.392	2.0	114	0.00
3	n-Nitrosodimethylamine	0.400	0.376	6.0	99	0.00
4 S	2-Fluorophenol	0.400	0.415	-3.7	110	0.00
5 S	Phenol-d6	0.400	0.431	-7.7	112	0.00
6	bis(2-Chloroethyl)ether	0.400	0.393	1.8	105	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	90	0.01
8 S	Nitrobenzene-d5	0.400	0.427	-6.7	107	0.01
9	Naphthalene	0.400	0.426	-6.5	106	0.01
10	Hexachlorobutadiene	0.400	0.406	-1.5	101	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.401	-0.3	101	0.01
12	2-Methylnaphthalene	0.400	0.421	-5.2	104	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.400	0.421	-5.2	103	0.00
15 S	2-Fluorobiphenyl	0.400	0.421	-5.2	102	0.00
16	Acenaphthylene	0.400	0.403	-0.8	96	0.01
17	Acenaphthene	0.400	0.410	-2.5	97	0.00
18	Fluorene	0.400	0.414	-3.5	97	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	85	0.00
20	4-Bromophenyl-phenylether	0.400	0.397	0.8	96	0.01
21	Hexachlorobenzene	0.400	0.411	-2.7	96	0.00
22	Pentachlorophenol	0.400	0.504	-26.0#	133	0.00
23	Phenanthrene	0.400	0.420	-5.0	99	0.00
24	Anthracene	0.400	0.412	-3.0	99	0.00
25 SURR	Fluoranthene-d10	0.400	0.403	-0.8	96	0.00
26	Fluoranthene	0.400	0.417	-4.2	98	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	88	0.00
28	Pvrene	0.400	0.450	-12.5	100	0.00
29 S	Terphenyl-d14	0.400	0.451	-12.7	100	0.00
30	Benzo(a)anthracene	0.400	0.458	-14.5	102	0.00
31	Chrysene	0.400	0.453	-13.2	99	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.455	-13.7	102	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.501	-25.2#	111	0.00
34 I	Perylene-d12	0.400	0.400	0.0	95	0.00
35	Benzo(b)fluoranthene	0.400	0.432	-8.0	112	0.00
36	Benzo(k)fluoranthene	0.400	0.418	-4.5	109	0.00
37 C	Benzo(a)pyrene	0.400	0.429	-7.2	112	0.00
38	Dibenzo(a,h)anthracene	0.400	0.435	-8.7	114	0.00
39	Benzo(a,h,i)perylene	0.400	0.429	-7.2	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE050819\
Data File : BE099579.D
Acq On : 8 May 2019 17:48
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 08 18:24:19 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 07 14:10:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				