

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE050919\
 Data File : BE099595.D
 Acq On : 9 May 2019 18:57
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 10 07:26:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 09 13:32:58 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	110	0.00
2	1,4-Dioxane	0.400	0.376	6.0	121	0.00
3	n-Nitrosodimethylamine	0.400	0.342	14.5	112	0.00
4 S	2-Fluorophenol	0.400	0.437	-9.2	138	0.00
5 S	Phenol-d6	0.400	0.422	-5.5	133	0.00
6	bis(2-Chloroethyl)ether	0.400	0.391	2.3	122	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	107	0.00
8 S	Nitrobenzene-d5	0.400	0.417	-4.2	126	0.00
9	Naphthalene	0.400	0.558	-39.5#	167	0.00
10	Hexachlorobutadiene	0.400	0.408	-2.0	121	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.444	-11.0	130	0.00
12	2-Methylnaphthalene	0.400	0.548	-37.0#	163	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.400	0.418	-4.5	128	0.00
15 S	2-Fluorobiphenyl	0.400	0.416	-4.0	124	0.00
16	Acenaphthylene	0.400	0.456	-14.0	137	0.00
17	Acenaphthene	0.400	0.408	-2.0	120	0.00
18	Fluorene	0.400	0.418	-4.5	126	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	105	0.00
20	4-Bromophenyl-phenylether	0.400	0.403	-0.8	121	0.00
21	Hexachlorobenzene	0.400	0.411	-2.7	119	0.00
22	Pentachlorophenol	0.400	0.549	-37.3#	149	0.00
23	Phenanthrene	0.400	0.597	-49.2#	175	0.00
24	Anthracene	0.400	0.446	-11.5	136	0.00
25 SURR	Fluoranthene-d10	0.400	0.431	-7.7	127	0.00
26	Fluoranthene	0.400	0.463	-15.8	136	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	109	0.00
28	Pvrene	0.400	0.524	-31.0#	144	0.00
29 S	Terphenyl-d14	0.400	0.450	-12.5	124	0.00
30	Benzo(a)anthracene	0.400	0.480	-20.0	131	0.00
31	Chrysene	0.400	0.465	-16.3	127	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.434	-8.5	123	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.454	-13.5	128	0.00
34 I	Perylene-d12	0.400	0.400	0.0	109	0.00
35	Benzo(b)fluoranthene	0.400	0.443	-10.7	133	0.00
36	Benzo(k)fluoranthene	0.400	0.426	-6.5	126	0.00
37 C	Benzo(a)pyrene	0.400	0.438	-9.5	132	0.00
38	Dibenzo(a,h)anthracene	0.400	0.417	-4.2	126	0.00
39	Benzo(a,h,i)perylene	0.400	0.415	-3.7	125	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE050919\
Data File : BE099595.D
Acq On : 9 May 2019 18:57
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 10 07:26:58 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 09 13:32:58 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				