

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE050919\
 Data File : BE099595.D
 Acq On : 9 May 2019 18:57
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 10 07:26:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 09 13:32:58 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.564	0.530	6.0	121	0.00
3	n-Nitrosodimethylamine	0.349	0.298	14.6	112	0.00
4 S	2-Fluorophenol	1.094	1.194	-9.1	138	0.00
5 S	Phenol-d6	1.437	1.518	-5.6	133	0.00
6	bis(2-Chloroethyl)ether	1.195	1.168	2.3	122	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
8 S	Nitrobenzene-d5	0.375	0.392	-4.5	126	0.00
9	Naphthalene	4.262	5.946	-39.5#	167#	0.00
10	Hexachlorobutadiene	0.308	0.314	-1.9	121	0.00
11 SURR	2-Methylnaphthalene-d10	0.670	0.744	-11.0	130	0.00
12	2-Methylnaphthalene	0.696	0.954	-37.1#	163#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.237	0.248	-4.6	128	0.00
15 S	2-Fluorobiphenyl	1.551	1.612	-3.9	124	0.00
16	Acenaphthylene	1.936	2.208	-14.0	137	0.00
17	Acenaphthene	1.081	1.102	-1.9	120	0.00
18	Fluorene	1.543	1.612	-4.5	126	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
20	4-Bromophenyl-phenylether	0.228	0.230	-0.9	121	0.00
21	Hexachlorobenzene	0.229	0.235	-2.6	119	0.00
22	Pentachlorophenol	0.092	0.096	-4.3	149	0.00
23	Phenanthrene	0.993	1.482	-49.2#	175#	0.00
24	Anthracene	0.932	1.039	-11.5	136	0.00
25 SURR	Fluoranthene-d10	5.261	5.667	-7.7	127	0.00
26	Fluoranthene	1.494	1.729	-15.7	136	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
28	Pvrene	1.054	1.381	-31.0#	144	0.00
29 S	Terphenyl-d14	0.711	0.799	-12.4	124	0.00
30	Benzo(a)anthracene	1.221	1.465	-20.0	131	0.00
31	Chrysene	1.221	1.421	-16.4	127	0.00
32	Bis(2-ethylhexyl)phthalate	1.755	1.907	-8.7	123	0.00
33	Indeno(1,2,3-cd)pyrene	1.524	1.728	-13.4	128	0.00
34 I	Perylene-d12	1.000	1.000	0.0	109	0.00
35	Benzo(b)fluoranthene	1.105	1.225	-10.9	133	0.00
36	Benzo(k)fluoranthene	1.098	1.169	-6.5	126	0.00
37 C	Benzo(a)pyrene	1.071	1.172	-9.4	132	0.00
38	Dibenzo(a,h)anthracene	1.119	1.166	-4.2	126	0.00
39	Benzo(a,h,i)perylene	1.133	1.176	-3.8	125	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE050919\
Data File : BE099595.D
Acq On : 9 May 2019 18:57
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 10 07:26:58 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 09 13:32:58 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			