

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE051019\
 Data File : BE099603.D
 Acq On : 10 May 2019 17:42
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 11 01:01:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 09 13:32:58 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	0.00
2	1,4-Dioxane	0.564	0.502	11.0	124	0.01
3	n-Nitrosodimethylamine	0.349	0.290	16.9	118	0.00
4 S	2-Fluorophenol	1.094	1.106	-1.1	139	0.00
5 S	Phenol-d6	1.437	1.441	-0.3	136	0.00
6	bis(2-Chloroethyl)ether	1.195	1.246	-4.3	141	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
8 S	Nitrobenzene-d5	0.375	0.387	-3.2	138	0.00
9	Naphthalene	4.262	4.444	-4.3	138	0.00
10	Hexachlorobutadiene	0.308	0.322	-4.5	137	0.00
11 SURR	2-Methylnaphthalene-d10	0.670	0.678	-1.2	131	0.00
12	2-Methylnaphthalene	0.696	0.718	-3.2	136	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
14 S	2,4,6-Tribromophenol	0.237	0.243	-2.5	140	0.00
15 S	2-Fluorobiphenyl	1.551	1.584	-2.1	136	0.00
16	Acenaphthylene	1.936	1.925	0.6	134	0.00
17	Acenaphthene	1.081	1.105	-2.2	135	0.00
18	Fluorene	1.543	1.592	-3.2	139	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
20	4-Bromophenyl-phenylether	0.228	0.233	-2.2	138	0.00
21	Hexachlorobenzene	0.229	0.244	-6.6	138	0.00
22	Pentachlorophenol	0.092	0.084	8.7	146	0.00
23	Phenanthrene	0.993	1.053	-6.0	140	0.00
24	Anthracene	0.932	0.954	-2.4	140	0.00
25 SURR	Fluoranthene-d10	5.261	5.292	-0.6	133	0.00
26	Fluoranthene	1.494	1.568	-5.0	138	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	118	0.00
28	Pvrene	1.054	1.206	-14.4	137	0.00
29 S	Terphenyl-d14	0.711	0.831	-16.9	140	0.00
30	Benzo(a)anthracene	1.221	1.406	-15.2	137	0.00
31	Chrysene	1.221	1.369	-12.1	133	0.00
32	Bis(2-ethylhexyl)phthalate	1.755	1.859	-5.9	131	0.00
33	Indeno(1,2,3-cd)pyrene	1.524	1.712	-12.3	138	0.00
34 I	Perylene-d12	1.000	1.000	0.0	119	0.00
35	Benzo(b)fluoranthene	1.105	1.168	-5.7	138	0.00
36	Benzo(k)fluoranthene	1.098	1.178	-7.3	138	0.00
37 C	Benzo(a)pyrene	1.071	1.138	-6.3	139	0.00
38	Dibenzo(a,h)anthracene	1.119	1.172	-4.7	138	0.00
39	Benzo(a,h,i)perylene	1.133	1.170	-3.3	135	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE051019\
Data File : BE099603.D
Acq On : 10 May 2019 17:42
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 11 01:01:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE050919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 09 13:32:58 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			