


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE051019\
 Data File : BE099596.D
 Acq On : 10 May 2019 9:26
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE050919.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu May 09 13:32:58 2019



AutoFind: Scans 2568, 2569, 2570; Background Corrected with Scan 2560

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	23.1	4705	PASS
68	69	0.00	2	1.3	86	PASS
69	198	0.00	100	33.4	6807	PASS
70	69	0.00	2	0.2	15	PASS
127	198	10	80	35.1	7170	PASS
197	198	0.00	2	0.2	32	PASS
198	198	100	100	100.0	20400	PASS
199	198	5	9	6.8	1391	PASS
275	198	10	60	30.7	6253	PASS
365	198	1	100	3.9	802	PASS
441	198	0.01	100	12.6	2578	PASS
442	198	50	100	81.3	16592	PASS
443	442	15	24	19.5	3242	PASS