

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE051419\
 Data File : BE099612.D
 Acq On : 14 May 2019 19:31
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE051419

Quant Time: May 15 05:02:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 19:30:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.590	0.586	0.7	96	0.00
3	n-Nitrosodimethylamine	0.376	0.369	1.9	120	0.00
4 S	2-Fluorophenol	1.136	1.143	-0.6	107	0.00
5 S	Phenol-d6	1.439	1.491	-3.6	113	0.00
6	bis(2-Chloroethyl)ether	1.211	1.239	-2.3	110	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.387	0.381	1.6	114	0.00
9	Naphthalene	4.256	4.300	-1.0	115	0.00
10	Hexachlorobutadiene	0.311	0.307	1.3	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.673	0.700	-4.0	117	0.00
12	2-Methylnaphthalene	0.703	0.698	0.7	112	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.274	0.247	9.9	102	0.00
15 S	2-Fluorobiphenyl	1.540	1.590	-3.2	119	0.00
16	Acenaphthylene	1.889	1.914	-1.3	117	0.00
17	Acenaphthene	1.066	1.092	-2.4	118	-0.01
18	Fluorene	1.529	1.544	-1.0	116	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
20	4-Bromophenyl-phenylether	0.231	0.227	1.7	115	0.00
21	Hexachlorobenzene	0.227	0.225	0.9	114	0.00
22	Pentachlorophenol	0.090	0.058	35.6#	98	0.01
23	Phenanthrene	0.993	0.985	0.8	115	0.00
24	Anthracene	0.916	0.892	2.6	116	0.00
25 SURR	Fluoranthene-d10	5.320	5.299	0.4	116	0.00
26	Fluoranthene	1.502	1.497	0.3	115	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
28	Pyrene	1.020	1.041	-2.1	117	0.00
29 S	Terphenyl-d14	0.720	0.718	0.3	116	0.00
30	Benzo(a)anthracene	1.209	1.188	1.7	114	0.00
31	Chrysene	1.245	1.265	-1.6	115	0.00
32	Bis(2-ethylhexyl)phthalate	1.864	1.746	6.3	112	0.00
33	Indeno(1,2,3-cd)pyrene	1.550	1.543	0.5	111	0.00
34 I	Perylene-d12	1.000	1.000	0.0	101	0.00
35	Benzo(b)fluoranthene	1.118	1.124	-0.5	113	0.00
36	Benzo(k)fluoranthene	1.105	1.122	-1.5	115	0.00
37 C	Benzo(a)pyrene	1.083	1.085	-0.2	112	0.00
38	Dibenzo(a,h)anthracene	1.123	1.112	1.0	110	0.00
39	Benzo(g,h,i)perylene	1.130	1.150	-1.8	112	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE051419\
Data File : BE099612.D
Acq On : 14 May 2019 19:31
Operator : JU/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE051419

Quant Time: May 15 05:02:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 19:30:50 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			