

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE051519\
 Data File : BE099614.D
 Acq On : 15 May 2019 9:12
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 15 23:40:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE051419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 19:30:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.590	0.554	6.1	101	0.00
3	n-Nitrosodimethylamine	0.376	0.305	18.9	111	0.00
4 S	2-Fluorophenol	1.136	1.015	10.7	106	0.00
5 S	Phenol-d6	1.439	1.301	9.6	110	0.00
6	bis(2-Chloroethyl)ether	1.211	1.097	9.4	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
8 S	Nitrobenzene-d5	0.387	0.340	12.1	111	0.00
9	Naphthalene	4.256	3.925	7.8	115	0.00
10	Hexachlorobutadiene	0.311	0.279	10.3	109	0.01
11 SURR	2-Methylnaphthalene-d10	0.673	0.611	9.2	112	0.00
12	2-Methylnaphthalene	0.703	0.635	9.7	111	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.274	0.217	20.8	102	0.00
15 S	2-Fluorobiphenyl	1.540	1.369	11.1	117	0.00
16	Acenaphthylene	1.889	1.666	11.8	116	0.00
17	Acenaphthene	1.066	0.953	10.6	117	0.00
18	Fluorene	1.529	1.380	9.7	118	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	119	0.00
20	4-Bromophenyl-phenylether	0.231	0.205	11.3	119	0.00
21	Hexachlorobenzene	0.227	0.209	7.9	122	0.00
22	Pentachlorophenol	0.090	0.038	57.8#	73	0.04
23	Phenanthrene	0.993	0.898	9.6	120	0.01
24	Anthracene	0.916	0.770	15.9	115	0.00
25 SURR	Fluoranthene-d10	5.320	4.871	8.4	123	0.00
26	Fluoranthene	1.502	1.374	8.5	121	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	126	0.00
28	Pvrene	1.020	1.007	1.3	126	0.00
29 S	Terphenyl-d14	0.720	0.681	5.4	123	0.00
30	Benzo(a)anthracene	1.209	1.137	6.0	122	0.00
31	Chrysene	1.245	1.229	1.3	124	0.00
32	Bis(2-ethylhexyl)phthalate	1.864	1.521	18.4	108	0.00
33	Indeno(1,2,3-cd)pyrene	1.550	1.515	2.3	121	0.00
34 I	Perylene-d12	1.000	1.000	0.0	120	0.00
35	Benzo(b)fluoranthene	1.118	0.999	10.6	120	0.00
36	Benzo(k)fluoranthene	1.105	1.011	8.5	123	0.00
37 C	Benzo(a)pyrene	1.083	0.981	9.4	121	0.00
38	Dibenzo(a,h)anthracene	1.123	1.009	10.2	119	0.00
39	Benzo(a,h,i)perylene	1.130	1.047	7.3	121	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE051519\
Data File : BE099614.D
Acq On : 15 May 2019 9:12
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 15 23:40:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE051419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 19:30:50 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			