

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE051519\
 Data File : BE099617.D
 Acq On : 15 May 2019 18:27
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 16 08:27:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE051419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 19:30:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	104	0.00
2	1,4-Dioxane	0.400	0.371	7.3	96	0.00
3	n-Nitrosodimethylamine	0.400	0.311	22.3	101	0.00
4 S	2-Fluorophenol	0.400	0.342	14.5	96	0.01
5 S	Phenol-d6	0.400	0.349	12.8	101	0.01
6	bis(2-Chloroethyl)ether	0.400	0.357	10.8	102	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	104	0.01
8 S	Nitrobenzene-d5	0.400	0.342	14.5	101	0.01
9	Naphthalene	0.400	0.367	8.3	106	0.01
10	Hexachlorobutadiene	0.400	0.359	10.3	102	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.364	9.0	104	0.01
12	2-Methylnaphthalene	0.400	0.353	11.8	101	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	102	0.01
14 S	2,4,6-Tribromophenol	0.400	0.323	19.3	92	0.00
15 S	2-Fluorobiphenyl	0.400	0.375	6.3	109	0.01
16	Acenaphthylene	0.400	0.357	10.8	104	0.00
17	Acenaphthene	0.400	0.365	8.8	106	0.00
18	Fluorene	0.400	0.349	12.8	101	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	0.01
20	4-Bromophenyl-phenylether	0.400	0.344	14.0	106	0.00
21	Hexachlorobenzene	0.400	0.352	12.0	107	0.01
22	Pentachlorophenol	0.400	0.431	-7.7	91	0.03
23	Phenanthrene	0.400	0.340	15.0	104	0.01
24	Anthracene	0.400	0.332	17.0	105	0.01
25 SURR	Fluoranthene-d10	0.400	0.350	12.5	108	0.00
26	Fluoranthene	0.400	0.353	11.8	108	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	118	0.01
28	Pvrene	0.400	0.371	7.3	111	0.00
29 S	Terphenyl-d14	0.400	0.357	10.8	108	0.00
30	Benzo(a)anthracene	0.400	0.402	-0.5	122	0.00
31	Chrysene	0.400	0.385	3.8	113	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.347	13.3	108	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.399	0.3	116	0.01
34 I	Perylene-d12	0.400	0.400	0.0	115	0.00
35	Benzo(b)fluoranthene	0.400	0.357	10.8	115	0.00
36	Benzo(k)fluoranthene	0.400	0.369	7.8	119	0.00
37 C	Benzo(a)pyrene	0.400	0.358	10.5	114	0.00
38	Dibenzo(a,h)anthracene	0.400	0.363	9.3	115	0.00
39	Benzo(a,h,i)perylene	0.400	0.370	7.5	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE051519\
Data File : BE099617.D
Acq On : 15 May 2019 18:27
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 16 08:27:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE051419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 19:30:50 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				