

Data Path : Z:\HPCHEM1\BNA_E\Data\BE051618\
 Data File : BE096587.D
 Acq On : 16 May 2018 8:41
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 17 03:54:32 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 15 13:02:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.600	0.585	2.5	111	0.00
3	n-Nitrosodimethylamine	0.750	0.740	1.3	109	0.00
4 S	2-Fluorophenol	1.103	1.164	-5.5	118	0.00
5 S	Phenol-d6	1.561	1.595	-2.2	109	0.00
6	bis(2-Chloroethyl)ether	1.460	1.501	-2.8	111	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.426	0.437	-2.6	112	0.00
9	Naphthalene	4.208	4.395	-4.4	112	0.00
10	Hexachlorobutadiene	0.235	0.221	6.0	110	0.00
11 SURR	2-Methylnaphthalene-d10	0.653	0.667	-2.1	110	0.00
12	2-Methylnaphthalene	0.721	0.739	-2.5	111	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.158	0.160	-1.3	109	0.00
15 S	2-Fluorobiphenyl	1.717	1.776	-3.4	111	0.00
16	Acenaphthylene	2.088	2.098	-0.5	108	0.00
17	Acenaphthene	1.272	1.271	0.1	106	0.00
18	Fluorene	1.583	1.615	-2.0	109	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.243	0.252	-3.7	109	0.00
21	Hexachlorobenzene	0.253	0.268	-5.9	111	0.00
22	Pentachlorophenol	0.053	0.044	17.0	128	0.00
23	Phenanthrene	1.123	1.150	-2.4	107	0.00
24	Anthracene	1.045	1.055	-1.0	106	0.00
25 SURR	Fluoranthene-d10	4.890	5.007	-2.4	108	0.00
26	Fluoranthene	1.343	1.359	-1.2	107	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pyrene	0.722	0.766	-6.1	111	0.00
29 S	Terphenyl-d14	0.905	0.956	-5.6	108	0.00
30	Benzo(a)anthracene	1.230	1.268	-3.1	109	0.00
31	Chrysene	1.280	1.315	-2.7	107	0.00
32	Bis(2-ethylhexyl)phthalate	2.632	2.697	-2.5	112	0.00
33	Indeno(1,2,3-cd)pyrene	1.227	1.267	-3.3	107	0.00
34 I	Perylene-d12	1.000	1.000	0.0	98	0.00
35	Benzo(b)fluoranthene	1.234	1.270	-2.9	108	0.00
36	Benzo(k)fluoranthene	1.306	1.323	-1.3	106	0.00
37 C	Benzo(a)pyrene	1.180	1.214	-2.9	107	0.00
38	Dibenzo(a,h)anthracene	1.079	1.110	-2.9	106	0.00
39	Benzo(g,h,i)perylene	1.129	1.172	-3.8	109	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE051618\
Data File : BE096587.D
Acq On : 16 May 2018 8:41
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 17 03:54:32 2018
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 15 13:02:03 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			