

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE051619\
 Data File : BE099633.D
 Acq On : 16 May 2019 21:28
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 17 07:39:35 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE051419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 19:30:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	125	0.00
2	1,4-Dioxane	0.400	0.323	19.3	100	0.00
3	n-Nitrosodimethylamine	0.400	0.327	18.3	128	0.01
4 S	2-Fluorophenol	0.400	0.346	13.5	117	0.00
5 S	Phenol-d6	0.400	0.364	9.0	127	0.00
6	bis(2-Chloroethyl)ether	0.400	0.355	11.3	123	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	126	0.00
8 S	Nitrobenzene-d5	0.400	0.349	12.8	125	0.00
9	Naphthalene	0.400	0.369	7.8	130	0.00
10	Hexachlorobutadiene	0.400	0.376	6.0	130	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.370	7.5	129	0.00
12	2-Methylnaphthalene	0.400	0.370	7.5	129	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	137	0.00
14 S	2,4,6-Tribromophenol	0.400	0.328	18.0	125	-0.01
15 S	2-Fluorobiphenyl	0.400	0.354	11.5	137	0.00
16	Acenaphthylene	0.400	0.349	12.8	136	0.00
17	Acenaphthene	0.400	0.356	11.0	138	-0.01
18	Fluorene	0.400	0.366	8.5	141	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	147	0.00
20	4-Bromophenyl-phenylether	0.400	0.358	10.5	148	0.00
21	Hexachlorobenzene	0.400	0.369	7.8	151	0.00
22	Pentachlorophenol	0.400	0.520	-30.0#	191	0.00
23	Phenanthrene	0.400	0.360	10.0	147	0.00
24	Anthracene	0.400	0.344	14.0	145	0.00
25 SURR	Fluoranthene-d10	0.400	0.359	10.3	148	0.00
26	Fluoranthene	0.400	0.365	8.8	149	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	140	0.00
28	Pvrene	0.400	0.426	-6.5	151	0.00
29 S	Terphenyl-d14	0.400	0.406	-1.5	146	0.00
30	Benzo(a)anthracene	0.400	0.392	2.0	141	0.00
31	Chrysene	0.400	0.405	-1.3	142	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.365	8.8	134	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.391	2.3	135	0.00
34 I	Perylene-d12	0.400	0.400	0.0	132	0.00
35	Benzo(b)fluoranthene	0.400	0.364	9.0	134	0.00
36	Benzo(k)fluoranthene	0.400	0.377	5.8	139	0.00
37 C	Benzo(a)pyrene	0.400	0.365	8.8	133	0.00
38	Dibenzo(a,h)anthracene	0.400	0.371	7.3	135	0.00
39	Benzo(a,h,i)perylene	0.400	0.374	6.5	134	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE051619\
Data File : BE099633.D
Acq On : 16 May 2019 21:28
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 17 07:39:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE051419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 19:30:50 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				