

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE051619\
 Data File : BE099633.D
 Acq On : 16 May 2019 21:28
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 17 07:39:35 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE051419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 19:30:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	0.00
2	1,4-Dioxane	0.590	0.476	19.3	100	0.00
3	n-Nitrosodimethylamine	0.376	0.307	18.4	128	0.01
4 S	2-Fluorophenol	1.136	0.981	13.6	117	0.00
5 S	Phenol-d6	1.439	1.310	9.0	127	0.00
6	bis(2-Chloroethyl)ether	1.211	1.074	11.3	123	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	126	0.00
8 S	Nitrobenzene-d5	0.387	0.338	12.7	125	0.00
9	Naphthalene	4.256	3.927	7.7	130	0.00
10	Hexachlorobutadiene	0.311	0.293	5.8	130	0.00
11 SURR	2-Methylnaphthalene-d10	0.673	0.623	7.4	129	0.00
12	2-Methylnaphthalene	0.703	0.650	7.5	129	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	137	0.00
14 S	2,4,6-Tribromophenol	0.274	0.225	17.9	125	-0.01
15 S	2-Fluorobiphenyl	1.540	1.363	11.5	137	0.00
16	Acenaphthylene	1.889	1.650	12.7	136	0.00
17	Acenaphthene	1.066	0.949	11.0	138	-0.01
18	Fluorene	1.529	1.398	8.6	141	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	147	0.00
20	4-Bromophenyl-phenylether	0.231	0.207	10.4	148	0.00
21	Hexachlorobenzene	0.227	0.209	7.9	151#	0.00
22	Pentachlorophenol	0.090	0.081	10.0	191#	0.00
23	Phenanthrene	0.993	0.893	10.1	147	0.00
24	Anthracene	0.916	0.787	14.1	145	0.00
25 SURR	Fluoranthene-d10	5.320	4.768	10.4	148	0.00
26	Fluoranthene	1.502	1.371	8.7	149	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	140	0.00
28	Pvrene	1.020	1.087	-6.6	151#	0.00
29 S	Terphenyl-d14	0.720	0.730	-1.4	146	0.00
30	Benzo(a)anthracene	1.209	1.183	2.2	141	0.00
31	Chrysene	1.245	1.260	-1.2	142	-0.01
32	Bis(2-ethylhexyl)phthalate	1.864	1.699	8.9	134	0.00
33	Indeno(1,2,3-cd)pyrene	1.550	1.516	2.2	135	0.00
34 I	Perylene-d12	1.000	1.000	0.0	132	0.00
35	Benzo(b)fluoranthene	1.118	1.017	9.0	134	0.00
36	Benzo(k)fluoranthene	1.105	1.042	5.7	139	0.00
37 C	Benzo(a)pyrene	1.083	0.988	8.8	133	0.00
38	Dibenzo(a,h)anthracene	1.123	1.042	7.2	135	0.00
39	Benzo(a,h,i)perylene	1.130	1.055	6.6	134	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE051619\
Data File : BE099633.D
Acq On : 16 May 2019 21:28
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 17 07:39:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE051419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 19:30:50 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			