

Data Path : Z:\HPCHEM1\BNA_E\Data\BE051718\
 Data File : BE096600.D
 Acq On : 17 May 2018 8:58
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 18 05:19:10 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 15 13:02:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.600	0.529	11.8	107	0.00
3	n-Nitrosodimethylamine	0.750	0.697	7.1	110	0.00
4 S	2-Fluorophenol	1.103	0.981	11.1	107	0.00
5 S	Phenol-d6	1.561	1.362	12.7	100	0.00
6	bis(2-Chloroethyl)ether	1.460	1.322	9.5	104	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.426	0.415	2.6	107	0.00
9	Naphthalene	4.208	4.053	3.7	104	0.00
10	Hexachlorobutadiene	0.235	0.226	3.8	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.653	0.621	4.9	104	0.00
12	2-Methylnaphthalene	0.721	0.686	4.9	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.158	0.137	13.3	95	0.00
15 S	2-Fluorobiphenyl	1.717	1.632	5.0	103	0.00
16	Acenaphthylene	2.088	1.944	6.9	102	0.00
17	Acenaphthene	1.272	1.213	4.6	102	-0.01
18	Fluorene	1.583	1.495	5.6	102	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
20	4-Bromophenyl-phenylether	0.243	0.235	3.3	103	0.00
21	Hexachlorobenzene	0.253	0.243	4.0	102	0.00
22	Pentachlorophenol	0.053	0.038	28.3#	110	0.00
23	Phenanthrene	1.123	1.049	6.6	98	0.00
24	Anthracene	1.045	0.967	7.5	98	0.00
25 SURR	Fluoranthene-d10	4.890	4.635	5.2	101	0.00
26	Fluoranthene	1.343	1.276	5.0	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
28	Pyrene	0.722	0.614	15.0	91	0.00
29 S	Terphenyl-d14	0.905	0.873	3.5	100	0.00
30	Benzo(a)anthracene	1.230	1.169	5.0	102	0.00
31	Chrysene	1.280	1.215	5.1	101	-0.01
32	Bis(2-ethylhexyl)phthalate	2.632	2.291	13.0	97	0.00
33	Indeno(1,2,3-cd)pyrene	1.227	1.150	6.3	99	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.234	1.169	5.3	101	0.00
36	Benzo(k)fluoranthene	1.306	1.236	5.4	102	0.00
37 C	Benzo(a)pyrene	1.180	1.127	4.5	102	0.00
38	Dibenzo(a,h)anthracene	1.079	0.997	7.6	98	0.00
39	Benzo(g,h,i)perylene	1.129	1.058	6.3	100	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE051718\
Data File : BE096600.D
Acq On : 17 May 2018 8:58
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 18 05:19:10 2018
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 15 13:02:03 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			