

Data Path : Z:\HPCHEM1\BNA_E\Data\BE051718\
 Data File : BE096615.D
 Acq On : 17 May 2018 18:30
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 18 05:53:25 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051518.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 15 13:02:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	106	0.00
2	1,4-Dioxane	0.400	0.378	5.5	112	0.00
3	n-Nitrosodimethylamine	0.400	0.372	7.0	108	0.00
4 S	2-Fluorophenol	0.400	0.363	9.3	106	0.00
5 S	Phenol-d6	0.400	0.360	10.0	101	0.00
6	bis(2-Chloroethyl)ether	0.400	0.385	3.8	108	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	107	-0.01
8 S	Nitrobenzene-d5	0.400	0.383	4.3	110	0.00
9	Naphthalene	0.400	0.374	6.5	106	0.00
10	Hexachlorobutadiene	0.400	0.361	9.8	111	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.366	8.5	104	0.00
12	2-Methylnaphthalene	0.400	0.367	8.3	105	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.400	0.364	9.0	98	0.00
15 S	2-Fluorobiphenyl	0.400	0.397	0.8	106	0.00
16	Acenaphthylene	0.400	0.384	4.0	104	0.00
17	Acenaphthene	0.400	0.382	4.5	101	-0.01
18	Fluorene	0.400	0.376	6.0	100	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.400	0.387	3.3	101	0.00
21	Hexachlorobenzene	0.400	0.386	3.5	101	0.00
22	Pentachlorophenol	0.400	0.368	8.0	109	0.00
23	Phenanthrene	0.400	0.378	5.5	98	0.00
24	Anthracene	0.400	0.368	8.0	96	0.00
25 SURR	Fluoranthene-d10	0.400	0.369	7.8	97	0.00
26	Fluoranthene	0.400	0.373	6.8	98	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	98	0.00
28	Pyrene	0.400	0.391	2.3	102	0.00
29 S	Terphenyl-d14	0.400	0.378	5.5	96	0.00
30	Benzo(a)anthracene	0.400	0.369	7.8	97	0.00
31	Chrysene	0.400	0.380	5.0	99	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.347	13.3	95	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.376	6.0	97	0.00
34 I	Perylene-d12	0.400	0.400	0.0	98	0.00
35	Benzo(b)fluoranthene	0.400	0.380	5.0	100	0.00
36	Benzo(k)fluoranthene	0.400	0.381	4.8	100	0.00
37 C	Benzo(a)pyrene	0.400	0.378	5.5	98	0.00
38	Dibenzo(a,h)anthracene	0.400	0.366	8.5	95	0.00
39	Benzo(g,h,i)perylene	0.400	0.378	5.5	99	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE051718\
Data File : BE096615.D
Acq On : 17 May 2018 18:30
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 18 05:53:25 2018
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051518.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 15 13:02:03 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				