


2,4,6-Tribromophenol

Response Ratio




Resp Ratio = $3.20 \times 10^{-1} \times \text{Amt} - 1.15 \times 10^{-1}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE052019.M
Calibration Table Last Updated: Mon May 20 19:29:40 2019

Pentachlorophenol

Response Ratio



$R = 4.16e-003 A^2 + 6.64e-002 A - 3.08e-002$
Coef of Det (r^2) = 1.000 Curve Fit: Quadratic

Method Name: Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE052019.M
Calibration Table Last Updated: Mon May 20 19:29:40 2019