

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052019\  
 Data File : BE099666.D  
 Acq On : 20 May 2019 18:37  
 Operator : JU/SJ  
 Sample : SSTDICV0.4  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 ICVBE052019

Quant Time: May 20 19:36:09 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon May 20 19:29:40 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.592	0.530	10.5	93	0.00
3	n-Nitrosodimethylamine	0.327	0.220	32.7#	76	0.00
4 S	2-Fluorophenol	1.011	0.866	14.3	97	0.00
5 S	Phenol-d6	1.338	1.163	13.1	104	0.00
6	bis(2-Chloroethyl)ether	1.190	1.035	13.0	98	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.346	0.294	15.0	97	0.00
9	Naphthalene	4.278	3.830	10.5	98	0.00
10	Hexachlorobutadiene	0.315	0.290	7.9	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.690	0.614	11.0	98	0.00
12	2-Methylnaphthalene	0.697	0.624	10.5	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.237	0.177	25.3#	96	0.00
15 S	2-Fluorobiphenyl	1.518	1.306	14.0	97	0.00
16	Acenaphthylene	1.864	1.602	14.1	100	0.00
17	Acenaphthene	1.046	0.908	13.2	98	0.01
18	Fluorene	1.541	1.353	12.2	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	4-Bromophenyl-phenylether	0.231	0.196	15.2	99	0.00
21	Hexachlorobenzene	0.236	0.207	12.3	100	0.00
22	Pentachlorophenol	0.068	0.028	58.8#	71	0.01
23	Phenanthrene	0.977	0.840	14.0	98	0.00
24	Anthracene	0.886	0.712	19.6	97	0.00
25 SURR	Fluoranthene-d10	5.466	4.720	13.6	100	0.00
26	Fluoranthene	1.545	1.353	12.4	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
28	Pyrene	1.005	0.966	3.9	100	0.00
29 S	Terphenyl-d14	0.714	0.670	6.2	99	0.00
30	Benzo(a)anthracene	1.137	1.044	8.2	98	0.00
31	Chrysene	1.215	1.187	2.3	101	0.01
32	Bis(2-ethylhexyl)phthalate	1.310	1.155	11.8	100	0.00
33	Indeno(1,2,3-cd)pyrene	1.467	1.398	4.7	103	0.00
34 I	Perylene-d12	1.000	1.000	0.0	102	0.00
35	Benzo(b)fluoranthene	1.096	0.978	10.8	101	0.00
36	Benzo(k)fluoranthene	1.143	1.007	11.9	100	0.00
37 C	Benzo(a)pyrene	1.055	0.952	9.8	104	0.00
38	Dibenzo(a,h)anthracene	1.108	0.965	12.9	103	0.00
39	Benzo(a,h,i)perylene	1.118	0.994	11.1	102	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052019\  
 Data File : BE099666.D  
 Acq On : 20 May 2019 18:37  
 Operator : JU/SJ  
 Sample : SSTDICV0.4  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

**Instrument :**  
 BNA\_E  
**ClientSampleId :**  
 ICVBE052019

Quant Time: May 20 19:36:09 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon May 20 19:29:40 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
-----				
(#) = Out of Range	SPCC's out = 0		CCC's out = 0	