

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052019\
 Data File : BE099668.D
 Acq On : 20 May 2019 19:54
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 21 13:43:01 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	0.00
2	1,4-Dioxane	0.400	0.359	10.3	89	0.00
3	n-Nitrosodimethylamine	0.400	0.311	22.3	84	0.00
4 S	2-Fluorophenol	0.400	0.349	12.8	95	0.00
5 S	Phenol-d6	0.400	0.336	16.0	96	0.00
6	bis(2-Chloroethyl)ether	0.400	0.348	13.0	94	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	0.00
8 S	Nitrobenzene-d5	0.400	0.335	16.3	94	0.00
9	Naphthalene	0.400	0.357	10.8	96	0.00
10	Hexachlorobutadiene	0.400	0.362	9.5	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.354	11.5	96	0.00
12	2-Methylnaphthalene	0.400	0.351	12.3	97	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.400	0.363	9.3	92	0.00
15 S	2-Fluorobiphenyl	0.400	0.360	10.0	98	0.00
16	Acenaphthylene	0.400	0.349	12.8	98	0.00
17	Acenaphthene	0.400	0.361	9.8	99	0.00
18	Fluorene	0.400	0.361	9.8	99	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	99	0.00
20	4-Bromophenyl-phenylether	0.400	0.340	15.0	97	0.00
21	Hexachlorobenzene	0.400	0.356	11.0	100	0.00
22	Pentachlorophenol	0.400	0.334	16.5	70	0.04
23	Phenanthrene	0.400	0.345	13.8	97	0.00
24	Anthracene	0.400	0.321	19.8	95	0.00
25 SURR	Fluoranthene-d10	0.400	0.329	17.8	94	0.00
26	Fluoranthene	0.400	0.332	17.0	94	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	95	0.00
28	Pvrene	0.400	0.392	2.0	95	0.00
29 S	Terphenyl-d14	0.400	0.381	4.8	94	0.00
30	Benzo(a)anthracene	0.400	0.371	7.3	93	0.00
31	Chrysene	0.400	0.392	2.0	94	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.324	19.0	86	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.381	4.8	96	0.00
34 I	Perylene-d12	0.400	0.400	0.0	95	0.00
35	Benzo(b)fluoranthene	0.400	0.354	11.5	94	0.00
36	Benzo(k)fluoranthene	0.400	0.364	9.0	97	0.00
37 C	Benzo(a)pyrene	0.400	0.359	10.3	97	0.00
38	Dibenzo(a,h)anthracene	0.400	0.349	12.8	96	0.00
39	Benzo(a,h,i)perylene	0.400	0.353	11.8	95	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052019\
Data File : BE099668.D
Acq On : 20 May 2019 19:54
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 21 13:43:01 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				