

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052019\
 Data File : BE099668.D
 Acq On : 20 May 2019 19:54
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 21 13:43:01 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.592	0.532	10.1	89	0.00
3	n-Nitrosodimethylamine	0.327	0.254	22.3	84	0.00
4 S	2-Fluorophenol	1.011	0.882	12.8	95	0.00
5 S	Phenol-d6	1.338	1.123	16.1	96	0.00
6	bis(2-Chloroethyl)ether	1.190	1.035	13.0	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.346	0.290	16.2	94	0.00
9	Naphthalene	4.278	3.814	10.8	96	0.00
10	Hexachlorobutadiene	0.315	0.285	9.5	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.690	0.611	11.4	96	0.00
12	2-Methylnaphthalene	0.697	0.612	12.2	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.237	0.175	26.2#	92	0.00
15 S	2-Fluorobiphenyl	1.518	1.365	10.1	98	0.00
16	Acenaphthylene	1.864	1.627	12.7	98	0.00
17	Acenaphthene	1.046	0.945	9.7	99	0.00
18	Fluorene	1.541	1.389	9.9	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4-Bromophenyl-phenylether	0.231	0.196	15.2	97	0.00
21	Hexachlorobenzene	0.236	0.210	11.0	100	0.00
22	Pentachlorophenol	0.068	0.028	58.8#	70	0.04
23	Phenanthrene	0.977	0.842	13.8	97	0.00
24	Anthracene	0.886	0.711	19.8	95	0.00
25 SURR	Fluoranthene-d10	5.466	4.491	17.8	94	0.00
26	Fluoranthene	1.545	1.284	16.9	94	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
28	Pvrene	1.005	0.985	2.0	95	0.00
29 S	Terphenyl-d14	0.714	0.681	4.6	94	0.00
30	Benzo(a)anthracene	1.137	1.053	7.4	93	0.00
31	Chrysene	1.215	1.192	1.9	94	0.00
32	Bis(2-ethylhexyl)phthalate	1.310	1.061	19.0	86	0.00
33	Indeno(1,2,3-cd)pyrene	1.467	1.396	4.8	96	0.00
34 I	Perylene-d12	1.000	1.000	0.0	95	0.00
35	Benzo(b)fluoranthene	1.096	0.970	11.5	94	0.00
36	Benzo(k)fluoranthene	1.143	1.039	9.1	97	0.00
37 C	Benzo(a)pyrene	1.055	0.946	10.3	97	0.00
38	Dibenzo(a,h)anthracene	1.108	0.965	12.9	96	0.00
39	Benzo(a,h,i)perylene	1.118	0.987	11.7	95	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052019\
Data File : BE099668.D
Acq On : 20 May 2019 19:54
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 21 13:43:01 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			