

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052019\
 Data File : BE099684.D
 Acq On : 21 May 2019 5:31
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: May 21 07:44:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2	1,4-Dioxane	0.592	0.535	9.6	87	0.00
3	n-Nitrosodimethylamine	0.327	0.210	35.8#	67	0.00
4 S	2-Fluorophenol	1.011	0.957	5.3	100	0.00
5 S	Phenol-d6	1.338	1.300	2.8	108	0.00
6	bis(2-Chloroethyl)ether	1.190	1.066	10.4	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.346	0.330	4.6	104	0.00
9	Naphthalene	4.278	17.758	-315.1#	435#	0.00
10	Hexachlorobutadiene	0.315	0.281	10.8	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.690	0.638	7.5	97	0.00
12	2-Methylnaphthalene	0.697	1.378	-97.7#	212#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.237	0.231	2.5	124	-0.01
15 S	2-Fluorobiphenyl	1.518	1.358	10.5	100	0.00
16	Acenaphthylene	1.864	1.889	-1.3	117	-0.01
17	Acenaphthene	1.046	0.974	6.9	105	0.00
18	Fluorene	1.541	1.943	-26.1#	143	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.231	0.208	10.0	101	0.00
21	Hexachlorobenzene	0.236	0.216	8.5	101	0.00
22	Pentachlorophenol	0.068	0.088	-29.4#	216#	-0.01
23	Phenanthrene	0.977	1.933	-97.9#	218#	0.00
24	Anthracene	0.886	1.021	-15.2	134	0.00
25 SURR	Fluoranthene-d10	5.466	4.916	10.1	100	0.00
26	Fluoranthene	1.545	1.917	-24.1	138	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
28	Pvrene	1.005	1.485	-47.8#	154#	0.00
29 S	Terphenyl-d14	0.714	0.691	3.2	103	0.00
30	Benzo(a)anthracene	1.137	1.409	-23.9	134	0.00
31	Chrysene	1.215	1.380	-13.6	118	0.00
32	Bis(2-ethylhexyl)phthalate	1.310	1.908	-45.6#	166#	0.00
33	Indeno(1,2,3-cd)pyrene	1.467	1.574	-7.3	117	0.00
34 I	Perylene-d12	1.000	1.000	0.0	107	0.00
35	Benzo(b)fluoranthene	1.096	1.147	-4.7	125	0.00
36	Benzo(k)fluoranthene	1.143	1.065	6.8	112	0.00
37 C	Benzo(a)pyrene	1.055	1.096	-3.9	126	0.00
38	Dibenzo(a,h)anthracene	1.108	1.010	8.8	113	0.00
39	Benzo(a,h,i)perylene	1.118	1.035	7.4	113	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052019\
Data File : BE099684.D
Acq On : 21 May 2019 5:31
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 21 07:44:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		