

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052019\
 Data File : BE099684.D
 Acq On : 21 May 2019 5:31
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 21 07:44:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	92	0.00
2	1,4-Dioxane	0.400	0.362	9.5	87	0.00
3	n-Nitrosodimethylamine	0.400	0.258	35.5#	67	0.00
4 S	2-Fluorophenol	0.400	0.379	5.3	100	0.00
5 S	Phenol-d6	0.400	0.389	2.8	108	0.00
6	bis(2-Chloroethyl)ether	0.400	0.358	10.5	94	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	94	0.00
8 S	Nitrobenzene-d5	0.400	0.382	4.5	104	0.00
9	Naphthalene	0.400	1.661	-315.3#	435	0.00
10	Hexachlorobutadiene	0.400	0.356	11.0	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.369	7.8	97	0.00
12	2-Methylnaphthalene	0.400	0.790	-97.5#	212	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.400	0.433	-8.2	124	-0.01
15 S	2-Fluorobiphenyl	0.400	0.358	10.5	100	0.00
16	Acenaphthylene	0.400	0.405	-1.3	117	-0.01
17	Acenaphthene	0.400	0.373	6.8	105	0.00
18	Fluorene	0.400	0.504	-26.0#	143	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.400	0.360	10.0	101	0.00
21	Hexachlorobenzene	0.400	0.366	8.5	101	0.00
22	Pentachlorophenol	0.400	0.647	-61.7#	216	-0.01
23	Phenanthrene	0.400	0.791	-97.8#	218	0.00
24	Anthracene	0.400	0.461	-15.3	134	0.00
25 SURR	Fluoranthene-d10	0.400	0.360	10.0	100	0.00
26	Fluoranthene	0.400	0.496	-24.0	138	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	103	0.00
28	Pvrene	0.400	0.591	-47.7#	154	0.00
29 S	Terphenyl-d14	0.400	0.387	3.3	103	0.00
30	Benzo(a)anthracene	0.400	0.496	-24.0	134	0.00
31	Chrysene	0.400	0.454	-13.5	118	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.583	-45.7#	166	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.429	-7.2	117	0.00
34 I	Perylene-d12	0.400	0.400	0.0	107	0.00
35	Benzo(b)fluoranthene	0.400	0.419	-4.7	125	0.00
36	Benzo(k)fluoranthene	0.400	0.373	6.8	112	0.00
37 C	Benzo(a)pyrene	0.400	0.416	-4.0	126	0.00
38	Dibenzo(a,h)anthracene	0.400	0.365	8.8	113	0.00
39	Benzo(a,h,i)perylene	0.400	0.370	7.5	113	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052019\
Data File : BE099684.D
Acq On : 21 May 2019 5:31
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 21 07:44:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				