

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052119\  
 Data File : BE099687.D  
 Acq On : 21 May 2019 13:09  
 Operator : JU/SJ  
 Sample : K2944-01  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 FESB-1(14.5-15)

Quant Time: May 21 14:09:54 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon May 20 19:29:40 2019  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc  | Units | Dev(Min)        |
|--------------------------------|-------|------|----------|-------|-------|-----------------|
| 1) 1,4-Dichlorobenzene-d4      | 7.81  | 152  | 1608     | 0.40  | ng    | 0.00            |
| 7) Naphthalene-d8              | 10.61 | 136  | 5478     | 0.40  | ng    | 0.00            |
| 13) Acenaphthene-d10           | 14.49 | 164  | 4061     | 0.40  | ng    | 0.00            |
| 19) Phenanthrene-d10           | 17.23 | 188  | 12015    | 0.40  | ng    | 0.00            |
| 27) Chrysene-d12               | 21.41 | 240  | 15518    | 0.40  | ng    | 0.00            |
| 34) Perylene-d12               | 23.95 | 264  | 18873    | 0.40  | ng    | 0.00            |
| System Monitoring Compounds    |       |      |          |       |       |                 |
| 4) 2-Fluorophenol              | 5.38  | 112  | 1497     | 0.37  | ng    | 0.00            |
| 5) Phenol-d6                   | 6.98  | 99   | 2030     | 0.38  | ng    | 0.00            |
| 8) Nitrobenzene-d5             | 8.98  | 82   | 1541     | 0.32  | ng    | 0.00            |
| 11) 2-Methylnaphthalene-d10    | 12.22 | 152  | 3304     | 0.35  | ng    | 0.00            |
| 14) 2,4,6-Tribromophenol       | 15.98 | 330  | 715      | 0.36  | ng    | 0.00            |
| 15) 2-Fluorobiphenyl           | 13.10 | 172  | 5036     | 0.33  | ng    | 0.00            |
| 25) Fluoranthene-d10           | 19.26 | 212  | 52068    | 0.32  | ng    | 0.00            |
| 29) Terphenyl-d14              | 19.86 | 244  | 11049    | 0.40  | ng    | 0.00            |
| Target Compounds               |       |      |          |       |       |                 |
| 32) Bis(2-ethylhexyl)phthalate | 21.32 | 149  | 14686    | 0.289 | ng    | Qvalue<br># 100 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052119\  
Data File : BE099687.D  
Acq On : 21 May 2019 13:09  
Operator : JU/SJ  
Sample : K2944-01  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
FESB-1(14.5-15)

Quant Time: May 21 14:09:54 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon May 20 19:29:40 2019  
Response via : Initial Calibration

