

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052119\
 Data File : BE099706.D
 Acq On : 22 May 2019 1:23
 Operator : JU/SJ
 Sample : K2906-01
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 FESB-11(7-7.5)

Quant Time: May 28 02:45:05 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.81	152	1431	0.40	ng	0.00
7) Naphthalene-d8	10.61	136	5090	0.40	ng	0.00
13) Acenaphthene-d10	14.49	164	3683	0.40	ng	0.00
19) Phenanthrene-d10	17.23	188	10843	0.40	ng	0.00
27) Chrysene-d12	21.41	240	14434	0.40	ng	0.00
34) Perylene-d12	23.96	264	17299	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.38	112	1443	0.40	ng	0.00
5) Phenol-d6	6.98	99	2064	0.43	ng	0.00
8) Nitrobenzene-d5	8.98	82	1723	0.39	ng	0.00
11) 2-Methylnaphthalene-d10	12.22	152	3239	0.37	ng	0.00
14) 2,4,6-Tribromophenol	15.97	330	799	0.42	ng	-0.01
15) 2-Fluorobiphenyl	13.10	172	4964	0.36	ng	0.00
25) Fluoranthene-d10	19.26	212	49092	0.33	ng	0.00
29) Terphenyl-d14	19.86	244	10367	0.40	ng	-0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
26) Fluoranthene	19.29	202	1218	0.029	ng	# 92
28) Pyrene	19.66	202	1173	0.032	ng	# 94
30) Benzo(a)anthracene	21.40	228	913	0.022	ng	# 64
32) Bis(2-ethylhexyl)phthalate	21.31	149	17337	0.367	ng	# 99
35) Benzo(b)fluoranthene	23.17	252	1920	0.041	ng	# 25
37) Benzo(a)pyrene	23.83	252	1155	0.025	ng	# 32
39) Benzo(g,h,i)perylene	27.48	276	1063	0.022	ng	# 64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052119\
Data File : BE099706.D
Acq On : 22 May 2019 1:23
Operator : JU/SJ
Sample : K2906-01
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
FESB-11(7-7.5)

Quant Time: May 28 02:45:05 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

