

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052119\
 Data File : BE099686.D
 Acq On : 21 May 2019 12:31
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 22 09:17:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	-0.01
2	1,4-Dioxane	0.592	0.478	19.3	81	0.00
3	n-Nitrosodimethylamine	0.327	0.305	6.7	102	0.00
4 S	2-Fluorophenol	1.011	0.935	7.5	102	0.00
5 S	Phenol-d6	1.338	1.256	6.1	108	0.00
6	bis(2-Chloroethyl)ether	1.190	1.091	8.3	100	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.346	0.312	9.8	105	0.00
9	Naphthalene	4.278	3.870	9.5	101	0.00
10	Hexachlorobutadiene	0.315	0.285	9.5	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.690	0.635	8.0	103	0.00
12	2-Methylnaphthalene	0.697	0.640	8.2	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.237	0.195	17.7	107	0.00
15 S	2-Fluorobiphenyl	1.518	1.362	10.3	102	0.00
16	Acenaphthylene	1.864	1.649	11.5	104	-0.01
17	Acenaphthene	1.046	0.976	6.7	107	0.00
18	Fluorene	1.541	1.363	11.6	102	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
20	4-Bromophenyl-phenylether	0.231	0.208	10.0	99	0.00
21	Hexachlorobenzene	0.236	0.224	5.1	102	0.00
22	Pentachlorophenol	0.068	0.054	20.6	131	0.00
23	Phenanthrene	0.977	0.881	9.8	97	0.00
24	Anthracene	0.886	0.802	9.5	103	0.00
25 SURR	Fluoranthene-d10	5.466	4.825	11.7	97	0.00
26	Fluoranthene	1.545	1.360	12.0	96	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
28	Pvrene	1.005	1.006	-0.1	98	0.00
29 S	Terphenyl-d14	0.714	0.682	4.5	96	0.00
30	Benzo(a)anthracene	1.137	1.142	-0.4	102	0.00
31	Chrysene	1.215	1.170	3.7	94	0.00
32	Bis(2-ethylhexyl)phthalate	1.310	1.543	-17.8	127	0.00
33	Indeno(1,2,3-cd)pyrene	1.467	1.380	5.9	97	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	96	0.00
35	Benzo(b)fluoranthene	1.096	1.003	8.5	98	0.00
36	Benzo(k)fluoranthene	1.143	1.010	11.6	95	0.00
37 C	Benzo(a)pyrene	1.055	0.948	10.1	98	0.00
38	Dibenzo(a,h)anthracene	1.108	0.975	12.0	98	0.00
39	Benzo(a,h,i)perylene	1.118	0.993	11.2	97	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052119\
Data File : BE099686.D
Acq On : 21 May 2019 12:31
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 22 09:17:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			