

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052119\
 Data File : BE099691.D
 Acq On : 21 May 2019 16:21
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 21 18:55:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.592	0.448	24.3	76	-0.01
3	n-Nitrosodimethylamine	0.327	0.223	31.8#	75	0.00
4 S	2-Fluorophenol	1.011	0.973	3.8	106	0.00
5 S	Phenol-d6	1.338	1.250	6.6	108	0.00
6	bis(2-Chloroethyl)ether	1.190	1.053	11.5	97	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.01
8 S	Nitrobenzene-d5	0.346	0.320	7.5	108	0.00
9	Naphthalene	4.278	3.736	12.7	98	0.00
10	Hexachlorobutadiene	0.315	0.274	13.0	98	0.01
11 SURR	2-Methylnaphthalene-d10	0.690	0.613	11.2	100	0.00
12	2-Methylnaphthalene	0.697	0.636	8.8	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.237	0.228	3.8	126	0.00
15 S	2-Fluorobiphenyl	1.518	1.322	12.9	100	0.01
16	Acenaphthylene	1.864	1.702	8.7	108	0.00
17	Acenaphthene	1.046	0.946	9.6	104	0.01
18	Fluorene	1.541	1.362	11.6	103	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4-Bromophenyl-phenylether	0.231	0.201	13.0	98	0.00
21	Hexachlorobenzene	0.236	0.218	7.6	102	0.00
22	Pentachlorophenol	0.068	0.057	16.2	142	0.00
23	Phenanthrene	0.977	0.897	8.2	101	0.01
24	Anthracene	0.886	0.792	10.6	104	0.00
25 SURR	Fluoranthene-d10	5.466	4.900	10.4	100	0.00
26	Fluoranthene	1.545	1.388	10.2	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	0.01
28	Pvrene	1.005	0.975	3.0	103	0.00
29 S	Terphenyl-d14	0.714	0.667	6.6	101	0.00
30	Benzo(a)anthracene	1.137	1.158	-1.8	111	0.00
31	Chrysene	1.215	1.216	-0.1	106	0.01
32	Bis(2-ethylhexyl)phthalate	1.310	1.676	-27.9#	148	0.00
33	Indeno(1,2,3-cd)pyrene	1.467	1.416	3.5	107	0.00
34 I	Perylene-d12	1.000	1.000	0.0	103	0.00
35	Benzo(b)fluoranthene	1.096	1.030	6.0	108	0.00
36	Benzo(k)fluoranthene	1.143	1.011	11.5	102	0.00
37 C	Benzo(a)pyrene	1.055	0.949	10.0	106	0.00
38	Dibenzo(a,h)anthracene	1.108	0.984	11.2	106	0.00
39	Benzo(a,h,i)perylene	1.118	0.988	11.6	104	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052119\
Data File : BE099691.D
Acq On : 21 May 2019 16:21
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 21 18:55:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			