

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052119\
 Data File : BE099693.D
 Acq On : 21 May 2019 17:31
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 21 18:59:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.592	0.497	16.0	89	0.00
3	n-Nitrosodimethylamine	0.327	0.186	43.1#	66	0.00
4 S	2-Fluorophenol	1.011	0.922	8.8	106	0.00
5 S	Phenol-d6	1.338	1.275	4.7	117	0.00
6	bis(2-Chloroethyl)ether	1.190	1.053	11.5	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.346	0.313	9.5	111	0.00
9	Naphthalene	4.278	3.867	9.6	107	0.00
10	Hexachlorobutadiene	0.315	0.287	8.9	108	0.00
11 SURR	2-Methylnaphthalene-d10	0.690	0.649	5.9	111	0.00
12	2-Methylnaphthalene	0.697	0.637	8.6	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
14 S	2,4,6-Tribromophenol	0.237	0.201	15.2	121	0.00
15 S	2-Fluorobiphenyl	1.518	1.323	12.8	109	0.00
16	Acenaphthylene	1.864	1.641	12.0	114	0.00
17	Acenaphthene	1.046	0.923	11.8	111	0.00
18	Fluorene	1.541	1.317	14.5	108	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	4-Bromophenyl-phenylether	0.231	0.209	9.5	105	0.00
21	Hexachlorobenzene	0.236	0.221	6.4	107	0.00
22	Pentachlorophenol	0.068	0.044	35.3#	114	0.01
23	Phenanthrene	0.977	0.875	10.4	102	0.00
24	Anthracene	0.886	0.777	12.3	106	0.00
25 SURR	Fluoranthene-d10	5.466	4.585	16.1	97	0.00
26	Fluoranthene	1.545	1.304	15.6	97	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
28	Pvrene	1.005	1.017	-1.2	99	0.00
29 S	Terphenyl-d14	0.714	0.684	4.2	96	0.00
30	Benzo(a)anthracene	1.137	1.154	-1.5	103	0.00
31	Chrysene	1.215	1.214	0.1	97	0.00
32	Bis(2-ethylhexyl)phthalate	1.310	1.591	-21.5	130	0.00
33	Indeno(1,2,3-cd)pyrene	1.467	1.428	2.7	100	0.00
34 I	Perylene-d12	1.000	1.000	0.0	98	0.00
35	Benzo(b)fluoranthene	1.096	1.011	7.8	100	0.00
36	Benzo(k)fluoranthene	1.143	1.019	10.8	97	0.00
37 C	Benzo(a)pyrene	1.055	0.949	10.0	100	0.00
38	Dibenzo(a,h)anthracene	1.108	0.969	12.5	99	0.00
39	Benzo(a,h,i)perylene	1.118	0.993	11.2	98	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052119\
Data File : BE099693.D
Acq On : 21 May 2019 17:31
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 21 18:59:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			