

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052119\
 Data File : BE099711.D
 Acq On : 22 May 2019 4:22
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: May 22 08:54:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.592	0.456	23.0	83	0.00
3	n-Nitrosodimethylamine	0.327	0.217	33.6#	78	0.00
4 S	2-Fluorophenol	1.011	0.969	4.2	113	0.00
5 S	Phenol-d6	1.338	1.301	2.8	121	0.00
6	bis(2-Chloroethyl)ether	1.190	1.048	11.9	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
8 S	Nitrobenzene-d5	0.346	0.312	9.8	117	0.00
9	Naphthalene	4.278	3.986	6.8	116	0.00
10	Hexachlorobutadiene	0.315	0.293	7.0	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.690	0.628	9.0	114	0.00
12	2-Methylnaphthalene	0.697	0.661	5.2	121	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
14 S	2,4,6-Tribromophenol	0.237	0.212	10.5	132	0.00
15 S	2-Fluorobiphenyl	1.518	1.345	11.4	114	0.00
16	Acenaphthylene	1.864	1.707	8.4	122	-0.02
17	Acenaphthene	1.046	0.970	7.3	120	0.00
18	Fluorene	1.541	1.382	10.3	117	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
20	4-Bromophenyl-phenylether	0.231	0.204	11.7	112	0.00
21	Hexachlorobenzene	0.236	0.214	9.3	113	0.00
22	Pentachlorophenol	0.068	0.051	25.0#	143	0.00
23	Phenanthrene	0.977	0.890	8.9	113	0.00
24	Anthracene	0.886	0.788	11.1	117	0.00
25 SURR	Fluoranthene-d10	5.466	4.532	17.1	105	0.00
26	Fluoranthene	1.545	1.294	16.2	105	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
28	Pvrene	1.005	1.004	0.1	108	0.00
29 S	Terphenyl-d14	0.714	0.671	6.0	104	0.00
30	Benzo(a)anthracene	1.137	1.141	-0.4	112	0.00
31	Chrysene	1.215	1.198	1.4	106	0.00
32	Bis(2-ethylhexyl)phthalate	1.310	1.471	-12.3	132	0.00
33	Indeno(1,2,3-cd)pyrene	1.467	1.449	1.2	112	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	106	0.00
35	Benzo(b)fluoranthene	1.096	1.009	7.9	109	0.00
36	Benzo(k)fluoranthene	1.143	1.020	10.8	106	0.00
37 C	Benzo(a)pyrene	1.055	0.958	9.2	109	0.00
38	Dibenzo(a,h)anthracene	1.108	1.008	9.0	112	0.00
39	Benzo(a,h,i)perylene	1.118	1.004	10.2	108	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052119\
Data File : BE099711.D
Acq On : 22 May 2019 4:22
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 22 08:54:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			