


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052119\
 Data File : BE099692.D
 Acq On : 21 May 2019 16:55
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE052019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon May 20 19:29:40 2019



AutoFind: Scans 2565, 2566, 2567; Background Corrected with Scan 2557

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	19.9	5832	PASS
68	69	0.00	2	1.7	148	PASS
69	198	0.00	100	29.5	8651	PASS
70	69	0.00	2	0.2	13	PASS
127	198	10	80	31.4	9208	PASS
197	198	0.00	2	0.3	85	PASS
198	198	100	100	100.0	29365	PASS
199	198	5	9	6.9	2029	PASS
275	198	10	60	31.3	9199	PASS
365	198	1	100	3.6	1065	PASS
441	198	0.01	100	14.2	4160	PASS
442	198	50	100	91.5	26872	PASS
443	442	15	24	19.0	5102	PASS