

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052219\
 Data File : BE099728.D
 Acq On : 22 May 2019 19:57
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: May 23 08:32:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	122	-0.01
2	1,4-Dioxane	0.592	0.462	22.0	100	0.00
3	n-Nitrosodimethylamine	0.327	0.234	28.4#	99	-0.01
4 S	2-Fluorophenol	1.011	1.041	-3.0	144	0.00
5 S	Phenol-d6	1.338	1.355	-1.3	149	0.00
6	bis(2-Chloroethyl)ether	1.190	1.074	9.7	126	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	123	0.00
8 S	Nitrobenzene-d5	0.346	0.346	0.0	144	-0.01
9	Naphthalene	4.278	4.071	4.8	132	0.00
10	Hexachlorobutadiene	0.315	0.302	4.1	133	0.00
11 SURR	2-Methylnaphthalene-d10	0.690	0.649	5.9	131	-0.01
12	2-Methylnaphthalene	0.697	0.665	4.6	135	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
14 S	2,4,6-Tribromophenol	0.237	0.254	-7.2	169#	-0.01
15 S	2-Fluorobiphenyl	1.518	1.413	6.9	128	0.00
16	Acenaphthylene	1.864	1.801	3.4	138	-0.01
17	Acenaphthene	1.046	1.006	3.8	133	0.00
18	Fluorene	1.541	1.415	8.2	129	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
20	4-Bromophenyl-phenylether	0.231	0.223	3.5	125	0.00
21	Hexachlorobenzene	0.236	0.230	2.5	124	0.00
22	Pentachlorophenol	0.068	0.093	-36.8#	267#	-0.01
23	Phenanthrene	0.977	0.945	3.3	123	0.00
24	Anthracene	0.886	0.863	2.6	131	0.00
25 SURR	Fluoranthene-d10	5.466	5.119	6.3	121	0.00
26	Fluoranthene	1.545	1.472	4.7	123	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	125	0.00
28	Pvrene	1.005	0.995	1.0	126	0.00
29 S	Terphenyl-d14	0.714	0.680	4.8	124	0.00
30	Benzo(a)anthracene	1.137	1.233	-8.4	143	-0.01
31	Chrysene	1.215	1.221	-0.5	128	0.00
32	Bis(2-ethylhexyl)phthalate	1.310	1.756	-34.0#	187#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.467	1.470	-0.2	134	0.00
34 I	Perylene-d12	1.000	1.000	0.0	123	0.00
35	Benzo(b)fluoranthene	1.096	1.061	3.2	132	0.00
36	Benzo(k)fluoranthene	1.143	1.010	11.6	121	0.00
37 C	Benzo(a)pyrene	1.055	1.010	4.3	133	0.00
38	Dibenzo(a,h)anthracene	1.108	1.055	4.8	135	0.00
39	Benzo(a,h,i)perylene	1.118	1.030	7.9	128	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052219\
Data File : BE099728.D
Acq On : 22 May 2019 19:57
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 23 08:32:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		