

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052319\
 Data File : BE099730.D
 Acq On : 23 May 2019 12:20
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 23 15:11:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	75	-0.01
2	1,4-Dioxane	0.400	0.411	-2.7	81	-0.01
3	n-Nitrosodimethylamine	0.400	0.453	-13.2	97	-0.02
4 S	2-Fluorophenol	0.400	0.463	-15.8	99	-0.01
5 S	Phenol-d6	0.400	0.445	-11.2	100	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.419	-4.7	90	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	73	-0.01
8 S	Nitrobenzene-d5	0.400	0.476	-19.0	101	-0.01
9	Naphthalene	0.400	0.405	-1.3	82	-0.01
10	Hexachlorobutadiene	0.400	0.417	-4.2	85	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.400	0.0	82	-0.02
12	2-Methylnaphthalene	0.400	0.406	-1.5	84	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	69	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.564	-41.0#	125	-0.01
15 S	2-Fluorobiphenyl	0.400	0.420	-5.0	81	0.00
16	Acenaphthylene	0.400	0.426	-6.5	85	-0.02
17	Acenaphthene	0.400	0.427	-6.7	83	0.00
18	Fluorene	0.400	0.423	-5.7	83	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	66	-0.01
20	4-Bromophenyl-phenylether	0.400	0.415	-3.7	79	-0.01
21	Hexachlorobenzene	0.400	0.399	0.3	74	-0.01
22	Pentachlorophenol	0.400	0.833	-108.2#	210	-0.03
23	Phenanthrene	0.400	0.408	-2.0	76	-0.01
24	Anthracene	0.400	0.418	-4.5	82	-0.01
25 SURR	Fluoranthene-d10	0.400	0.409	-2.2	77	-0.01
26	Fluoranthene	0.400	0.404	-1.0	76	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	78	-0.01
28	Pvrene	0.400	0.395	1.3	78	-0.02
29 S	Terphenyl-d14	0.400	0.395	1.3	80	-0.02
30	Benzo(a)anthracene	0.400	0.437	-9.2	90	-0.01
31	Chrysene	0.400	0.412	-3.0	81	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.582	-45.5#	126	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.441	-10.2	92	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	77	-0.02
35	Benzo(b)fluoranthene	0.400	0.403	-0.8	86	-0.02
36	Benzo(k)fluoranthene	0.400	0.391	2.3	84	-0.02
37 C	Benzo(a)pyrene	0.400	0.411	-2.7	89	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.412	-3.0	92	-0.03
39	Benzo(a,h,i)perylene	0.400	0.409	-2.2	89	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052319\
Data File : BE099730.D
Acq On : 23 May 2019 12:20
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 23 15:11:53 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				