

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052319\  
 Data File : BE099734.D  
 Acq On : 23 May 2019 15:43  
 Operator : JU/SJ  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: May 23 16:27:40 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon May 20 19:29:40 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|         | Compound                   | AvqRF | CCRF  | %Dev   | Area% | Dev(min) |
|---------|----------------------------|-------|-------|--------|-------|----------|
| 1 I     | 1,4-Dichlorobenzene-d4     | 1.000 | 1.000 | 0.0    | 77    | 0.00     |
| 2       | 1,4-Dioxane                | 0.592 | 0.566 | 4.4    | 77    | -0.01    |
| 3       | n-Nitrosodimethylamine     | 0.327 | 0.368 | -12.5  | 99    | -0.02    |
| 4 S     | 2-Fluorophenol             | 1.011 | 1.134 | -12.2  | 99    | 0.00     |
| 5 S     | Phenol-d6                  | 1.338 | 1.496 | -11.8  | 104   | 0.00     |
| 6       | bis(2-Chloroethyl)ether    | 1.190 | 1.226 | -3.0   | 91    | 0.00     |
| 7 I     | Naphthalene-d8             | 1.000 | 1.000 | 0.0    | 75    | 0.01     |
| 8 S     | Nitrobenzene-d5            | 0.346 | 0.408 | -17.9  | 103   | 0.00     |
| 9       | Naphthalene                | 4.278 | 4.331 | -1.2   | 85    | 0.01     |
| 10      | Hexachlorobutadiene        | 0.315 | 0.320 | -1.6   | 85    | 0.01     |
| 11 SURR | 2-Methylnaphthalene-d10    | 0.690 | 0.716 | -3.8   | 87    | 0.00     |
| 12      | 2-Methylnaphthalene        | 0.697 | 0.713 | -2.3   | 88    | 0.00     |
| 13 I    | Acenaphthene-d10           | 1.000 | 1.000 | 0.0    | 71    | 0.00     |
| 14 S    | 2,4,6-Tribromophenol       | 0.237 | 0.307 | -29.5# | 118   | -0.01    |
| 15 S    | 2-Fluorobiphenyl           | 1.518 | 1.622 | -6.9   | 85    | 0.01     |
| 16      | Acenaphthylene             | 1.864 | 2.013 | -8.0   | 89    | 0.00     |
| 17      | Acenaphthene               | 1.046 | 1.130 | -8.0   | 86    | 0.00     |
| 18      | Fluorene                   | 1.541 | 1.587 | -3.0   | 83    | 0.00     |
| 19 I    | Phenanthrene-d10           | 1.000 | 1.000 | 0.0    | 67    | 0.00     |
| 20      | 4-Bromophenyl-phenylether  | 0.231 | 0.234 | -1.3   | 79    | 0.00     |
| 21      | Hexachlorobenzene          | 0.236 | 0.242 | -2.5   | 78    | 0.00     |
| 22      | Pentachlorophenol          | 0.068 | 0.125 | -83.8# | 214#  | -0.01    |
| 23      | Phenanthrene               | 0.977 | 0.990 | -1.3   | 77    | 0.00     |
| 24      | Anthracene                 | 0.886 | 0.930 | -5.0   | 85    | 0.00     |
| 25 SURR | Fluoranthene-d10           | 5.466 | 5.542 | -1.4   | 78    | 0.00     |
| 26      | Fluoranthene               | 1.545 | 1.559 | -0.9   | 78    | 0.00     |
| 27 I    | Chrysene-d12               | 1.000 | 1.000 | 0.0    | 79    | 0.00     |
| 28      | Pvrene                     | 1.005 | 1.005 | 0.0    | 80    | 0.00     |
| 29 S    | Terphenyl-d14              | 0.714 | 0.705 | 1.3    | 81    | -0.01    |
| 30      | Benzo(a)anthracene         | 1.137 | 1.275 | -12.1  | 93    | -0.01    |
| 31      | Chrysene                   | 1.215 | 1.222 | -0.6   | 80    | 0.00     |
| 32      | Bis(2-ethylhexyl)phthalate | 1.310 | 1.825 | -39.3# | 122   | -0.01    |
| 33      | Indeno(1,2,3-cd)pyrene     | 1.467 | 1.572 | -7.2   | 90    | 0.00     |
| 34 I    | Perylene-d12               | 1.000 | 1.000 | 0.0    | 76    | 0.00     |
| 35      | Benzo(b)fluoranthene       | 1.096 | 1.134 | -3.5   | 88    | 0.00     |
| 36      | Benzo(k)fluoranthene       | 1.143 | 1.107 | 3.1    | 83    | 0.00     |
| 37 C    | Benzo(a)pyrene             | 1.055 | 1.086 | -2.9   | 89    | 0.00     |
| 38      | Dibenzo(a,h)anthracene     | 1.108 | 1.131 | -2.1   | 90    | -0.01    |
| 39      | Benzo(a,h,i)perylene       | 1.118 | 1.134 | -1.4   | 88    | 0.00     |

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052319\  
Data File : BE099734.D  
Acq On : 23 May 2019 15:43  
Operator : JU/SJ  
Sample : SSTDCCC0.4  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_E  
LabSampleId :  
SSTDCCC0.4

Quant Time: May 23 16:27:40 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon May 20 19:29:40 2019  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound           | AvgRF                        | CCRF | %Dev Area | % Dev(min) |
|--------------------|------------------------------|------|-----------|------------|
| -----              |                              |      |           |            |
| (#) = Out of Range | SPCC's out = 0 CCC's out = 0 |      |           |            |