


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052319\
 Data File : BE099729.D
 Acq On : 23 May 2019 11:42
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE052019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon May 20 19:29:40 2019



AutoFind: Scans 2563, 2564, 2565; Background Corrected with Scan 2555

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	20.9	4523	PASS
68	69	0.00	2	1.6	103	PASS
69	198	0.00	100	30.5	6615	PASS
70	69	0.00	2	0.6	39	PASS
127	198	10	80	31.5	6841	PASS
197	198	0.00	2	0.3	74	PASS
198	198	100	100	100.0	21690	PASS
199	198	5	9	7.2	1562	PASS
275	198	10	60	30.7	6665	PASS
365	198	1	100	3.7	795	PASS
441	198	0.01	100	13.4	2899	PASS
442	198	50	100	86.6	18778	PASS
443	442	15	24	20.9	3922	PASS