

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052419\
 Data File : BE099740.D
 Acq On : 24 May 2019 13:47
 Operator : JU/SJ
 Sample : K3010-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 FESB-20(2.5-3)

Quant Time: May 24 21:46:37 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	1190	0.40	ng	-0.01
7) Naphthalene-d8	10.61	136	4028	0.40	ng	0.00
13) Acenaphthene-d10	14.47	164	3028	0.40	ng	-0.02
19) Phenanthrene-d10	17.23	188	9049	0.40	ng	0.00
27) Chrysene-d12	21.40	240	14060	0.40	ng	-0.01
34) Perylene-d12	23.94	264	17102	0.40	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.36	112	1514	0.50	ng	-0.01
5) Phenol-d6	6.96	99	2103	0.53	ng	-0.01
8) Nitrobenzene-d5	8.96	82	1750	0.50	ng	-0.01
11) 2-Methylnaphthalene-d10	12.21	152	3091	0.44	ng	-0.02
14) 2,4,6-Tribromophenol	15.97	330	854	0.50	ng	-0.01
15) 2-Fluorobiphenyl	13.10	172	4861	0.42	ng	0.00
25) Fluoranthene-d10	19.25	212	54498	0.44	ng	-0.01
29) Terphenyl-d14	19.85	244	11412	0.45	ng	-0.02
Target Compounds						Qvalue
32) Bis(2-ethylhexyl)phthalate	21.31	149	5328	0.116	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052419\
Data File : BE099740.D
Acq On : 24 May 2019 13:47
Operator : JU/SJ
Sample : K3010-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
FESB-20(2.5-3)

Quant Time: May 24 21:46:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

