

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052419\
 Data File : BE099736.D
 Acq On : 24 May 2019 11:23
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 24 21:37:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	60	-0.01
2	1,4-Dioxane	0.592	0.640	-8.1	68	-0.02
3	n-Nitrosodimethylamine	0.327	0.380	-16.2	80	-0.03
4 S	2-Fluorophenol	1.011	1.144	-13.2	78	-0.01
5 S	Phenol-d6	1.338	1.511	-12.9	82	-0.01
6	bis(2-Chloroethyl)ether	1.190	1.288	-8.2	75	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	63	0.00
8 S	Nitrobenzene-d5	0.346	0.400	-15.6	85	-0.01
9	Naphthalene	4.278	4.399	-2.8	73	-0.01
10	Hexachlorobutadiene	0.315	0.323	-2.5	73	0.00
11 SURR	2-Methylnaphthalene-d10	0.690	0.717	-3.9	74	-0.02
12	2-Methylnaphthalene	0.697	0.742	-6.5	77	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	65	0.00
14 S	2,4,6-Tribromophenol	0.237	0.317	-33.8#	112	-0.01
15 S	2-Fluorobiphenyl	1.518	1.565	-3.1	75	0.00
16	Acenaphthylene	1.864	1.970	-5.7	80	-0.02
17	Acenaphthene	1.046	1.089	-4.1	76	0.00
18	Fluorene	1.541	1.643	-6.6	79	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	68	0.00
20	4-Bromophenyl-phenylether	0.231	0.231	0.0	78	-0.01
21	Hexachlorobenzene	0.236	0.236	0.0	77	-0.01
22	Pentachlorophenol	0.068	0.102	-50.0#	176#	-0.03
23	Phenanthrene	0.977	1.034	-5.8	81	0.00
24	Anthracene	0.886	0.930	-5.0	85	0.00
25 SURR	Fluoranthene-d10	5.466	5.814	-6.4	83	0.00
26	Fluoranthene	1.545	1.640	-6.1	82	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
28	Pvrene	1.005	1.022	-1.7	86	-0.01
29 S	Terphenyl-d14	0.714	0.715	-0.1	86	-0.01
30	Benzo(a)anthracene	1.137	1.207	-6.2	92	-0.01
31	Chrysene	1.215	1.248	-2.7	86	0.00
32	Bis(2-ethylhexyl)phthalate	1.310	1.765	-34.7#	124	-0.01
33	Indeno(1,2,3-cd)pyrene	1.467	1.491	-1.6	90	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	76	-0.02
35	Benzo(b)fluoranthene	1.096	1.122	-2.4	86	-0.01
36	Benzo(k)fluoranthene	1.143	1.137	0.5	84	-0.02
37 C	Benzo(a)pyrene	1.055	1.087	-3.0	89	-0.01
38	Dibenzo(a,h)anthracene	1.108	1.111	-0.3	88	-0.02
39	Benzo(a,h,i)perylene	1.118	1.123	-0.4	86	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052419\
Data File : BE099736.D
Acq On : 24 May 2019 11:23
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 24 21:37:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			