

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052419\
 Data File : BE099753.D
 Acq On : 24 May 2019 21:35
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 24 22:10:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	64	-0.01
2	1,4-Dioxane	0.400	0.378	5.5	63	-0.02
3	n-Nitrosodimethylamine	0.400	0.412	-3.0	75	-0.02
4 S	2-Fluorophenol	0.400	0.430	-7.5	79	-0.01
5 S	Phenol-d6	0.400	0.432	-8.0	83	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.413	-3.2	75	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	67	-0.01
8 S	Nitrobenzene-d5	0.400	0.427	-6.7	83	-0.01
9	Naphthalene	0.400	0.405	-1.3	76	-0.01
10	Hexachlorobutadiene	0.400	0.392	2.0	73	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.422	-5.5	79	-0.02
12	2-Methylnaphthalene	0.400	0.402	-0.5	77	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	72	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.504	-26.0#	112	-0.01
15 S	2-Fluorobiphenyl	0.400	0.386	3.5	78	0.00
16	Acenaphthylene	0.400	0.412	-3.0	86	-0.02
17	Acenaphthene	0.400	0.417	-4.2	85	0.00
18	Fluorene	0.400	0.431	-7.7	88	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	77	-0.01
20	4-Bromophenyl-phenylether	0.400	0.396	1.0	87	-0.01
21	Hexachlorobenzene	0.400	0.388	3.0	84	-0.01
22	Pentachlorophenol	0.400	0.594	-48.5#	150	-0.03
23	Phenanthrene	0.400	0.404	-1.0	88	-0.01
24	Anthracene	0.400	0.404	-1.0	93	-0.01
25 SURR	Fluoranthene-d10	0.400	0.413	-3.2	91	-0.01
26	Fluoranthene	0.400	0.414	-3.5	90	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	84	-0.01
28	Pvrene	0.400	0.432	-8.0	93	-0.02
29 S	Terphenyl-d14	0.400	0.408	-2.0	89	-0.02
30	Benzo(a)anthracene	0.400	0.418	-4.5	92	-0.01
31	Chrysene	0.400	0.414	-3.5	88	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.504	-26.0#	118	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.405	-1.3	91	-0.03
34 I	Perylene-d12	0.400	0.400	0.0	78	-0.02
35	Benzo(b)fluoranthene	0.400	0.410	-2.5	89	-0.02
36	Benzo(k)fluoranthene	0.400	0.405	-1.3	88	-0.02
37 C	Benzo(a)pyrene	0.400	0.412	-3.0	91	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.401	-0.3	90	-0.03
39	Benzo(a,h,i)perylene	0.400	0.404	-1.0	89	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052419\
Data File : BE099753.D
Acq On : 24 May 2019 21:35
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 24 22:10:58 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				