

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052419\
 Data File : BE099753.D
 Acq On : 24 May 2019 21:35
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 24 22:10:58 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	64	-0.01
2	1,4-Dioxane	0.592	0.559	5.6	63	-0.02
3	n-Nitrosodimethylamine	0.327	0.337	-3.1	75	-0.02
4 S	2-Fluorophenol	1.011	1.088	-7.6	79	-0.01
5 S	Phenol-d6	1.338	1.444	-7.9	83	-0.01
6	bis(2-Chloroethyl)ether	1.190	1.229	-3.3	75	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	67	-0.01
8 S	Nitrobenzene-d5	0.346	0.369	-6.6	83	-0.01
9	Naphthalene	4.278	4.333	-1.3	76	-0.01
10	Hexachlorobutadiene	0.315	0.308	2.2	73	-0.01
11 SURR	2-Methylnaphthalene-d10	0.690	0.729	-5.7	79	-0.02
12	2-Methylnaphthalene	0.697	0.700	-0.4	77	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	72	-0.02
14 S	2,4,6-Tribromophenol	0.237	0.288	-21.5	112	-0.01
15 S	2-Fluorobiphenyl	1.518	1.463	3.6	78	0.00
16	Acenaphthylene	1.864	1.920	-3.0	86	-0.02
17	Acenaphthene	1.046	1.092	-4.4	85	0.00
18	Fluorene	1.541	1.662	-7.9	88	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	77	-0.01
20	4-Bromophenyl-phenylether	0.231	0.228	1.3	87	-0.01
21	Hexachlorobenzene	0.236	0.229	3.0	84	-0.01
22	Pentachlorophenol	0.068	0.077	-13.2	150	-0.03
23	Phenanthrene	0.977	0.988	-1.1	88	-0.01
24	Anthracene	0.886	0.894	-0.9	93	-0.01
25 SURR	Fluoranthene-d10	5.466	5.638	-3.1	91	-0.01
26	Fluoranthene	1.545	1.598	-3.4	90	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	84	-0.01
28	Pvrene	1.005	1.087	-8.2	93	-0.02
29 S	Terphenyl-d14	0.714	0.729	-2.1	89	-0.02
30	Benzo(a)anthracene	1.137	1.188	-4.5	92	-0.01
31	Chrysene	1.215	1.259	-3.6	88	-0.01
32	Bis(2-ethylhexyl)phthalate	1.310	1.649	-25.9#	118	-0.01
33	Indeno(1,2,3-cd)pyrene	1.467	1.487	-1.4	91	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	78	-0.02
35	Benzo(b)fluoranthene	1.096	1.125	-2.6	89	-0.02
36	Benzo(k)fluoranthene	1.143	1.156	-1.1	88	-0.02
37 C	Benzo(a)pyrene	1.055	1.088	-3.1	91	-0.02
38	Dibenzo(a,h)anthracene	1.108	1.111	-0.3	90	-0.03
39	Benzo(a,h,i)perylene	1.118	1.130	-1.1	89	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052419\
Data File : BE099753.D
Acq On : 24 May 2019 21:35
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 24 22:10:58 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		