

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052919\
 Data File : BE099772.D
 Acq On : 29 May 2019 12:32
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: May 29 17:25:15 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.592	0.549	7.3	84	-0.02
3	n-Nitrosodimethylamine	0.327	0.291	11.0	88	-0.02
4 S	2-Fluorophenol	1.011	0.931	7.9	91	0.00
5 S	Phenol-d6	1.338	1.200	10.3	94	0.00
6	bis(2-Chloroethyl)ether	1.190	1.029	13.5	86	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
8 S	Nitrobenzene-d5	0.346	0.332	4.0	95	-0.01
9	Naphthalene	4.278	3.906	8.7	86	-0.01
10	Hexachlorobutadiene	0.315	0.288	8.6	87	0.00
11 SURR	2-Methylnaphthalene-d10	0.690	0.621	10.0	85	-0.02
12	2-Methylnaphthalene	0.697	0.623	10.6	87	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
14 S	2,4,6-Tribromophenol	0.237	0.206	13.1	96	-0.01
15 S	2-Fluorobiphenyl	1.518	1.340	11.7	85	0.00
16	Acenaphthylene	1.864	1.673	10.2	90	-0.02
17	Acenaphthene	1.046	0.919	12.1	85	0.00
18	Fluorene	1.541	1.357	11.9	86	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
20	4-Bromophenyl-phenylether	0.231	0.198	14.3	82	-0.01
21	Hexachlorobenzene	0.236	0.201	14.8	80	-0.01
22	Pentachlorophenol	0.068	0.043	36.8#	90	-0.01
23	Phenanthrene	0.977	0.862	11.8	83	0.00
24	Anthracene	0.886	0.762	14.0	85	0.00
25 SURR	Fluoranthene-d10	5.466	4.815	11.9	84	-0.01
26	Fluoranthene	1.545	1.373	11.1	84	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
28	Pvrene	1.005	0.969	3.6	86	-0.01
29 S	Terphenyl-d14	0.714	0.653	8.5	83	-0.01
30	Benzo(a)anthracene	1.137	1.159	-1.9	94	-0.01
31	Chrysene	1.215	1.141	6.1	83	0.00
32	Bis(2-ethylhexyl)phthalate	1.310	1.467	-12.0	109	-0.01
33	Indeno(1,2,3-cd)pyrene	1.467	1.415	3.5	90	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	87	-0.01
35	Benzo(b)fluoranthene	1.096	1.024	6.6	91	-0.01
36	Benzo(k)fluoranthene	1.143	0.996	12.9	85	-0.01
37 C	Benzo(a)pyrene	1.055	0.948	10.1	89	-0.01
38	Dibenzo(a,h)anthracene	1.108	0.988	10.8	90	-0.02
39	Benzo(a,h,i)perylene	1.118	1.011	9.6	89	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052919\
Data File : BE099772.D
Acq On : 29 May 2019 12:32
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 29 17:25:15 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			