

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052919\
 Data File : BE099782.D
 Acq On : 29 May 2019 18:38
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 30 01:34:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 20 19:29:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	-0.01
2	1,4-Dioxane	0.592	0.478	19.3	83	-0.01
3	n-Nitrosodimethylamine	0.327	0.320	2.1	109	-0.01
4 S	2-Fluorophenol	1.011	0.988	2.3	109	0.00
5 S	Phenol-d6	1.338	1.269	5.2	112	-0.01
6	bis(2-Chloroethyl)ether	1.190	1.090	8.4	102	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	101	-0.01
8 S	Nitrobenzene-d5	0.346	0.323	6.6	109	-0.01
9	Naphthalene	4.278	3.808	11.0	100	-0.01
10	Hexachlorobutadiene	0.315	0.286	9.2	103	-0.01
11 SURR	2-Methylnaphthalene-d10	0.690	0.621	10.0	102	-0.01
12	2-Methylnaphthalene	0.697	0.620	11.0	102	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	-0.01
14 S	2,4,6-Tribromophenol	0.237	0.210	11.4	115	-0.01
15 S	2-Fluorobiphenyl	1.518	1.315	13.4	98	0.00
16	Acenaphthylene	1.864	1.726	7.4	109	-0.01
17	Acenaphthene	1.046	0.962	8.0	105	0.00
18	Fluorene	1.541	1.379	10.5	103	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	-0.01
20	4-Bromophenyl-phenylether	0.231	0.201	13.0	100	-0.01
21	Hexachlorobenzene	0.236	0.207	12.3	99	-0.01
22	Pentachlorophenol	0.068	0.054	20.6	137	-0.01
23	Phenanthrene	0.977	0.868	11.2	100	0.00
24	Anthracene	0.886	0.772	12.9	104	0.00
25 SURR	Fluoranthene-d10	5.466	4.597	15.9	96	-0.01
26	Fluoranthene	1.545	1.306	15.5	96	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	-0.01
28	Pvrene	1.005	1.017	-1.2	98	-0.01
29 S	Terphenyl-d14	0.714	0.683	4.3	95	-0.02
30	Benzo(a)anthracene	1.137	1.109	2.5	98	-0.01
31	Chrysene	1.215	1.192	1.9	95	-0.01
32	Bis(2-ethylhexyl)phthalate	1.310	1.606	-22.6	130	-0.01
33	Indeno(1,2,3-cd)pyrene	1.467	1.390	5.2	96	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	93	-0.02
35	Benzo(b)fluoranthene	1.096	1.021	6.8	97	-0.01
36	Benzo(k)fluoranthene	1.143	0.981	14.2	90	-0.02
37 C	Benzo(a)pyrene	1.055	0.955	9.5	96	-0.02
38	Dibenzo(a,h)anthracene	1.108	0.966	12.8	94	-0.02
39	Benzo(a,h,i)perylene	1.118	0.993	11.2	94	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE052919\
Data File : BE099782.D
Acq On : 29 May 2019 18:38
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 30 01:34:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE052019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 20 19:29:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			